

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061623\
 Data File : PP058530.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Jun 2023 17:48
 Operator : YP\AJ
 Sample : PB153485BL
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB153485BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 16 22:14:38 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061523.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 16 05:48:25 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.457	3.766	55576408	36877717	25.050	23.197
2) SA Decachlor...	10.356	8.864	38489952	34314966	28.347	23.475

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061623\
Data File : PP058530.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jun 2023 17:48
Operator : YP\AJ
Sample : PB153485BL
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB153485BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 16 22:14:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061523.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 16 05:48:25 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.457 min
Delta R.T.: 0.000 min
Response: 55576408
Conc: 25.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB153485BL




#1 Tetrachloro-m-xylene

R.T.: 3.766 min
Delta R.T.: -0.002 min
Response: 36877717
Conc: 23.20 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.356 min
Delta R.T.: 0.003 min
Response: 38489952
Conc: 28.35 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.864 min
Delta R.T.: -0.010 min
Response: 34314966
Conc: 23.48 ng/ml