

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061721\
 Data File : PP036502.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 17 Jun 2021 18:35
 Operator : DD\AJ
 Sample : AR1660ICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 18 02:12:18 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 18 02:10:15 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.964	3.958	247420	164683	5.768	5.394
2) SA Decachlor...	10.998	9.262	250838	185304	5.970	5.899
Target Compounds						
3) L1 AR-1016-1	6.284	5.213	99006	68314	60.386	59.317
4) L1 AR-1016-2	6.308	5.233	139973	91561	60.141	57.879
5) L1 AR-1016-3	6.375	5.423	88818	51187	59.434	57.876
6) L1 AR-1016-4	6.482	5.477	73503	41839	58.217	60.440
7) L1 AR-1016-5	6.796	5.705	74816	52066	60.250	58.738
31) L7 AR-1260-1	7.971	6.803	135050	95568	65.975	60.950
32) L7 AR-1260-2	8.238	7.002	159196	113221	66.133	59.936
33) L7 AR-1260-3	8.603	7.156	121477	106317	62.985	60.097
34) L7 AR-1260-4	8.834	7.640	142785	87824	62.861	57.739
35) L7 AR-1260-5	9.162	7.888	267887	205560	62.525	56.822



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061721\
Data File : PP036502.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 17 Jun 2021 18:35
Operator : DD\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 18 02:12:18 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061721.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 18 02:10:15 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.964 min
Delta R.T.: -0.002 min
Response: 247420
Conc: 5.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.958 min
Delta R.T.: -0.002 min
Response: 164683
Conc: 5.39 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.998 min
Delta R.T.: -0.004 min
Response: 250838
Conc: 5.97 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.262 min
Delta R.T.: -0.003 min
Response: 185304
Conc: 5.90 ng/ml



#3 AR-1016-1


R.T.: 6.284 min
Delta R.T.: -0.002 min
Response: 99006
Conc: 60.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#3 AR-1016-1

R.T.: 5.213 min
Delta R.T.: -0.002 min
Response: 68314
Conc: 59.32 ng/ml




#4 AR-1016-2

R.T.: 6.308 min
Delta R.T.: -0.002 min
Response: 139973
Conc: 60.14 ng/ml



#4 AR-1016-2


R.T.: 5.233 min
Delta R.T.: -0.001 min
Response: 91561
Conc: 57.88 ng/ml



#5 AR-1016-3


R.T.: 6.375 min
Delta R.T.: -0.002 min
Response: 88818
Conc: 59.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#5 AR-1016-3

R.T.: 5.423 min
Delta R.T.: -0.002 min
Response: 51187
Conc: 57.88 ng/ml




#6 AR-1016-4

R.T.: 6.482 min
Delta R.T.: -0.002 min
Response: 73503
Conc: 58.22 ng/ml



#6 AR-1016-4


R.T.: 5.477 min
Delta R.T.: -0.001 min
Response: 41839
Conc: 60.44 ng/ml



#7 AR-1016-5


R.T.: 6.796 min
Delta R.T.: -0.002 min
Response: 74816
Conc: 60.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#7 AR-1016-5

R.T.: 5.705 min
Delta R.T.: -0.001 min
Response: 52066
Conc: 58.74 ng/ml




#31 AR-1260-1

R.T.: 7.971 min
Delta R.T.: -0.002 min
Response: 135050
Conc: 65.98 ng/ml



#31 AR-1260-1


R.T.: 6.803 min
Delta R.T.: 0.000 min
Response: 95568
Conc: 60.95 ng/ml



#32 AR-1260-2


R.T.: 8.238 min
Delta R.T.: -0.002 min
Response: 159196
Conc: 66.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#32 AR-1260-2

R.T.: 7.002 min
Delta R.T.: 0.000 min
Response: 113221
Conc: 59.94 ng/ml




#33 AR-1260-3

R.T.: 8.603 min
Delta R.T.: -0.002 min
Response: 121477
Conc: 62.98 ng/ml



#33 AR-1260-3


R.T.: 7.156 min
Delta R.T.: 0.000 min
Response: 106317
Conc: 60.10 ng/ml



#34 AR-1260-4


R.T.: 8.834 min
Delta R.T.: -0.002 min
Response: 142785
Conc: 62.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050




#34 AR-1260-4

R.T.: 7.640 min
Delta R.T.: -0.001 min
Response: 87824
Conc: 57.74 ng/ml



#35 AR-1260-5

R.T.: 9.162 min
Delta R.T.: -0.003 min
Response: 267887
Conc: 62.52 ng/ml



#35 AR-1260-5

R.T.: 7.888 min
Delta R.T.: -0.002 min
Response: 205560
Conc: 56.82 ng/ml