

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061722\
 Data File : PP048900.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jun 2022 17:33
 Operator : YP\AJ
 Sample : N3383-06
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 RB-04-(3-5)

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/20/2022
 Supervised By :Ankita Jodhani 06/20/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 17 18:43:21 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 07 11:09:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.033	3.198	45722596	44149192	19.081m	23.196
2) SA Decachlor...	10.384	8.604	38233967	24263590	20.060	19.814m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061722\
Data File : PP048900.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jun 2022 17:33
Operator : YP\AJ
Sample : N3383-06
Misc :
ALS Vial : 19 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
RB-04-(3-5)


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/20/2022
Supervised By :Ankita Jodhani 06/20/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 17 18:43:21 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 07 11:09:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.033 min
Delta R.T.: -0.002 min
Response: 45722596
Conc: 19.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
RB-04-(3-5)


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/20/2022
Supervised By :Ankita Jodhani 06/20/2022




#1 Tetrachloro-m-xylene

R.T.: 3.198 min
Delta R.T.: -0.007 min
Response: 44149192
Conc: 23.20 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.384 min
Delta R.T.: -0.003 min
Response: 38233967
Conc: 20.06 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.604 min
Delta R.T.: -0.011 min
Response: 24263590
Conc: 19.81 ng/ml m