

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061722\
 Data File : PP048919.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jun 2022 01:15
 Operator : YP\AJ
 Sample : N3371-01
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 VNJ-211

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/20/2022
 Supervised By :Ankita Jodhani 06/20/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 18 03:31:26 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 07 11:09:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.034	3.197	58313359	54307744	24.336	28.533
2) SA Decachlor...	10.381	8.602	46380508	31835403	24.334	25.997m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061722\
Data File : PP048919.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jun 2022 01:15
Operator : YP\AJ
Sample : N3371-01
Misc :
ALS Vial : 34 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
VNJ-211


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/20/2022
Supervised By :Ankita Jodhani 06/20/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 18 03:31:26 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 07 11:09:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 4.034 min
Delta R.T.: -0.001 min
Response: 58313359
Conc: 24.34 ng/ml

Instrument :

ECD_P


ClientSampleId :

VNJ-211

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 06/20/2022
Supervised By :Ankita Jodhani 06/20/2022




#1 Tetrachloro-m-xylene

R.T.: 3.197 min
Delta R.T.: -0.008 min
Response: 54307744
Conc: 28.53 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.381 min
Delta R.T.: -0.006 min
Response: 46380508
Conc: 24.33 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.602 min
Delta R.T.: -0.013 min
Response: 31835403
Conc: 26.00 ng/ml m