

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061724\
 Data File : PP065736.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jun 2024 19:13
 Operator : YP\AJ
 Sample : P2885-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 CL-02-061424

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/18/2024
 Supervised By :Ankita Jodhani 06/18/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 18 00:02:18 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 06 05:25:59 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.828	4.068	34575959	49844930	18.227	17.851
2) SA Decachlor...	10.843	9.252	34869478	23560576	17.728	15.993
Target Compounds						
26) L6 AR-1254-1	6.867	5.998	2985803	8680270	38.114	70.447 #
27) L6 AR-1254-2	7.086	6.144	7033032	4098595	59.536	38.926 #
28) L6 AR-1254-3	7.456	6.558	6697667	8844402	54.884	55.891m
29) L6 AR-1254-4	7.737	6.783	5054010	2675180	56.833	28.972 #
30) L6 AR-1254-5	8.158	7.204	7781938	10376341	79.441m	81.071

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061724\
Data File : PP065736.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jun 2024 19:13
Operator : YP\AJ
Sample : P2885-01
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
CL-02-061424

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/18/2024
Supervised By :Ankita Jodhani 06/18/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 18 00:02:18 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 06 05:25:59 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

