

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061725\
 Data File : PP073017.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jun 2025 19:53
 Operator : YP\AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 18 04:53:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061725.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 18 04:47:35 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.496	3.787	58356921	50461433	28.236	27.486
2) SA Decachlor...	10.188	8.798	89712688	66943275	27.993	27.448
Target Compounds						
41) L9 AR-1268-1	8.701	7.681	99649369	78256106	279.995	278.690
42) L9 AR-1268-2	8.794	7.748	85229522	70075104	281.317	278.402
43) L9 AR-1268-3	9.023	7.951	73808064	57897140	281.523	273.672
44) L9 AR-1268-4	9.440	8.244	32217924	25871226	272.895	278.171
45) L9 AR-1268-5	9.852	8.541	205.0E6	155.9E6	277.502	265.410



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061725\
Data File : PP073017.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jun 2025 19:53
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 18 04:53:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061725.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 18 04:47:35 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.496 min
Delta R.T.: 0.004 min
Response: 58356921
Conc: 28.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.787 min
Delta R.T.: 0.000 min
Response: 50461433
Conc: 27.49 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.188 min
Delta R.T.: 0.004 min
Response: 89712688
Conc: 27.99 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.798 min
Delta R.T.: 0.000 min
Response: 66943275
Conc: 27.45 ng/ml



#41 AR-1268-1


R.T.: 8.701 min
Delta R.T.: 0.003 min
Response: 99649369
Conc: 280.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#41 AR-1268-1

R.T.: 7.681 min
Delta R.T.: 0.000 min
Response: 78256106
Conc: 278.69 ng/ml




#42 AR-1268-2

R.T.: 8.794 min
Delta R.T.: 0.004 min
Response: 85229522
Conc: 281.32 ng/ml



#42 AR-1268-2


R.T.: 7.748 min
Delta R.T.: 0.000 min
Response: 70075104
Conc: 278.40 ng/ml



#43 AR-1268-3


R.T.: 9.023 min
Delta R.T.: 0.004 min
Response: 73808064
Conc: 281.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250




#43 AR-1268-3

R.T.: 7.951 min
Delta R.T.: 0.000 min
Response: 57897140
Conc: 273.67 ng/ml




#44 AR-1268-4

R.T.: 9.440 min
Delta R.T.: 0.004 min
Response: 32217924
Conc: 272.89 ng/ml



#44 AR-1268-4


R.T.: 8.244 min
Delta R.T.: 0.000 min
Response: 25871226
Conc: 278.17 ng/ml



#45 AR-1268-5

R.T.: 9.852 min
Delta R.T.: 0.004 min
Response: 204983110
Conc: 277.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250



#45 AR-1268-5

R.T.: 8.541 min
Delta R.T.: 0.000 min
Response: 155898408
Conc: 265.41 ng/ml