

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061821\
 Data File : PP036554.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Jun 2021 14:02
 Operator : DD\AJ
 Sample : PB137194BL
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 19 04:29:29 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 18 06:06:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.964	3.958	954095	665272	21.587	21.479
2)	SA Decachlor...	10.998	9.261	1043499	721610	24.325	22.564
Target Compounds							
8)	L2 AR-1221-1	5.216	0.000	22649	0	40.342	N.D. #
9)	L2 AR-1221-2	5.317	4.307	13366	9437	31.016	31.615
28)	L6 AR-1254-3	7.821f	0.000	22664	0	7.652	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061821\
Data File : PP036554.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Jun 2021 14:02
Operator : DD\AJ
Sample : PB137194BL
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 19 04:29:29 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061721.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 18 06:06:19 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.964 min
Delta R.T.: 0.000 min
Response: 954095
Conc: 21.59 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.958 min
Delta R.T.: -0.001 min
Response: 665272
Conc: 21.48 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.998 min
Delta R.T.: 0.000 min
Response: 1043499
Conc: 24.33 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.261 min
Delta R.T.: -0.002 min
Response: 721610
Conc: 22.56 ng/ml



#8 AR-1221-1


R.T.: 5.216 min
Delta R.T.: 0.000 min
Response: 22649
Conc: 40.34 ng/ml

Instrument :
ECD_P
ClientSampleId :




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.217 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 5.317 min
Delta R.T.: 0.001 min
Response: 13366
Conc: 31.02 ng/ml



#9 AR-1221-2


R.T.: 4.307 min
Delta R.T.: -0.009 min
Response: 9437
Conc: 31.62 ng/ml



#28 AR-1254-3

R.T.: 7.821 min
Delta R.T.: 0.016 min
Response: 22664
Conc: 7.65 ng/ml

Instrument :
ECD_P
ClientSampleId :



#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.665 min
Response: 0
Conc: N.D.