

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061924\
 Data File : PP065820.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jun 2024 15:58
 Operator : YP\AJ
 Sample : P2938-01
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 397

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/20/2024
 Supervised By :Ankita Jodhani 06/20/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 19 16:10:46 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 06 05:25:59 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.830	4.074	24519581	34641136	12.926	12.406m
2) SA Decachlor...	10.836	9.245	37271539	25944859	18.950	17.611
Target Compounds						
31) L7 AR-1260-1	7.623	6.694	26419423	24009328	272.770	226.248
32) L7 AR-1260-2	7.875	6.878	33170583	28112183	295.849	227.950
33) L7 AR-1260-3	8.238	7.036	25162237	25724680	288.203	221.811
34) L7 AR-1260-4	8.481	7.507	29192075	20863930	291.306	236.887
35) L7 AR-1260-5	8.826	7.744	59170335	46264301	313.759	241.701

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061924\
Data File : PP065820.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jun 2024 15:58
Operator : YP\AJ
Sample : P2938-01
Misc :
ALS Vial : 17 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
397


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/20/2024
Supervised By :Ankita Jodhani 06/20/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 19 16:10:46 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 06 05:25:59 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.830 min
Delta R.T.: 0.004 min
Response: 24519581
Conc: 12.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
397


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/20/2024
Supervised By :Ankita Jodhani 06/20/2024




#1 Tetrachloro-m-xylene

R.T.: 4.074 min
Delta R.T.: 0.003 min
Response: 34641136
Conc: 12.41 ng/ml m




#2 Decachlorobiphenyl

R.T.: 10.836 min
Delta R.T.: -0.006 min
Response: 37271539
Conc: 18.95 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.245 min
Delta R.T.: -0.015 min
Response: 25944859
Conc: 17.61 ng/ml




#31 AR-1260-1

R.T.: 7.623 min
Delta R.T.: 0.008 min
Response: 26419423
Conc: 272.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
397


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/20/2024
Supervised By :Ankita Jodhani 06/20/2024




#31 AR-1260-1

R.T.: 6.694 min
Delta R.T.: 0.000 min
Response: 24009328
Conc: 226.25 ng/ml




#32 AR-1260-2

R.T.: 7.875 min
Delta R.T.: 0.006 min
Response: 33170583
Conc: 295.85 ng/ml



#32 AR-1260-2

R.T.: 6.878 min
Delta R.T.: 0.000 min
Response: 28112183
Conc: 227.95 ng/ml




#33 AR-1260-3

R.T.: 8.238 min
Delta R.T.: 0.003 min
Response: 25162237
Conc: 288.20 ng/ml

Instrument :
ECD_P
ClientSampleId :
397


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/20/2024
Supervised By :Ankita Jodhani 06/20/2024




#33 AR-1260-3

R.T.: 7.036 min
Delta R.T.: -0.002 min
Response: 25724680
Conc: 221.81 ng/ml




#34 AR-1260-4

R.T.: 8.481 min
Delta R.T.: 0.001 min
Response: 29192075
Conc: 291.31 ng/ml



#34 AR-1260-4

R.T.: 7.507 min
Delta R.T.: -0.006 min
Response: 20863930
Conc: 236.89 ng/ml




#35 AR-1260-5

R.T.: 8.826 min
Delta R.T.: 0.000 min
Response: 59170335
Conc: 313.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
397

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/20/2024
Supervised By :Ankita Jodhani 06/20/2024



#35 AR-1260-5

R.T.: 7.744 min
Delta R.T.: -0.007 min
Response: 46264301
Conc: 241.70 ng/ml