

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061924\
 Data File : PP065861.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jun 2024 06:20
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/20/2024
 Supervised By :Ankita Jodhani 06/20/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 20 06:32:51 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 06 05:25:59 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.826	4.065	93789222	137.7E6	49.443	49.306
2)	SA Decachlor...	10.836	9.242	100.7E6	73317697	51.181	49.768
Target Compounds							
26)	L6 AR-1254-1	6.863	5.993	38752537	55996844	494.684	454.457
27)	L6 AR-1254-2	7.081	6.139	56691102	47109310	479.898	447.416
28)	L6 AR-1254-3	7.449	6.552	60503343	70817345	495.794	447.521
29)	L6 AR-1254-4	7.734	6.778	45046175	40895875	506.554	442.895m
30)	L6 AR-1254-5	8.154	7.201	49669334	55207615	507.047m	431.342

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP061924\
Data File : PP065861.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2024 06:20
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/20/2024
Supervised By :Ankita Jodhani 06/20/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 20 06:32:51 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 06 05:25:59 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.826 min
Delta R.T.: 0.000 min
Response: 93789222
Conc: 49.44 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/20/2024
Supervised By :Ankita Jodhani 06/20/2024




#1 Tetrachloro-m-xylene

R.T.: 4.065 min
Delta R.T.: -0.006 min
Response: 137678990
Conc: 49.31 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.836 min
Delta R.T.: -0.007 min
Response: 100665800
Conc: 51.18 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.242 min
Delta R.T.: -0.018 min
Response: 73317697
Conc: 49.77 ng/ml




#26 AR-1254-1

R.T.: 6.863 min
Delta R.T.: -0.001 min
Response: 38752537
Conc: 494.68 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/20/2024
Supervised By :Ankita Jodhani 06/20/2024




#26 AR-1254-1

R.T.: 5.993 min
Delta R.T.: -0.010 min
Response: 55996844
Conc: 454.46 ng/ml




#27 AR-1254-2

R.T.: 7.081 min
Delta R.T.: -0.001 min
Response: 56691102
Conc: 479.90 ng/ml



#27 AR-1254-2

R.T.: 6.139 min
Delta R.T.: -0.010 min
Response: 47109310
Conc: 447.42 ng/ml




#28 AR-1254-3

R.T.: 7.449 min
Delta R.T.: -0.003 min
Response: 60503343
Conc: 495.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/20/2024
Supervised By :Ankita Jodhani 06/20/2024




#28 AR-1254-3

R.T.: 6.552 min
Delta R.T.: -0.010 min
Response: 70817345
Conc: 447.52 ng/ml




#29 AR-1254-4

R.T.: 7.734 min
Delta R.T.: -0.002 min
Response: 45046175
Conc: 506.55 ng/ml



#29 AR-1254-4

R.T.: 6.778 min
Delta R.T.: -0.011 min
Response: 40895875
Conc: 442.90 ng/ml m




#30 AR-1254-5

R.T.: 8.154 min
Delta R.T.: -0.003 min
Response: 49669334
Conc: 507.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/20/2024
Supervised By :Ankita Jodhani 06/20/2024



#30 AR-1254-5

R.T.: 7.201 min
Delta R.T.: -0.011 min
Response: 55207615
Conc: 431.34 ng/ml