

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062018\
 Data File : PP026531.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Jun 2018 14:42
 Operator : UA/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 20 14:54:34 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062018.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 20 14:06:28 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-CLPEST1 Signal #2 Phase: ZB-CLPEST2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.063	4.136	4950671	12437063	53.656	52.230
2)	SA Decachlor...	8.335	8.953	7938652	8020880	54.720	50.835
Target Compounds							
3)	L1 Aroclor-1...	4.570	4.490	1247552	2037176	543.480	501.014
4)	L1 Aroclor-1...	4.995	4.960	1423341	1544138	537.850	512.280
5)	L1 Aroclor-1...	5.041	5.202	1003262	3162459	552.040	514.314
6)	L1 Aroclor-1...	5.351	5.281	675020	2054714	553.859	517.397
7)	L1 Aroclor-1...	5.377	5.545	780027	2284886	515.407	550.470
31)	L7 Aroclor-1...	6.193	6.514	1825674	4153973	512.243	505.244
32)	L7 Aroclor-1...	6.629	6.666	4701604	4619124	513.073	503.052
33)	L7 Aroclor-1...	6.715	7.035	3300701	3279348	500.415	500.434
34)	L7 Aroclor-1...	6.915	7.287	2721447	3708275	525.235	505.330
35)	L7 Aroclor-1...	7.161	7.480	9104596	8057293	493.595	509.521



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062018\
Data File : PP026531.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Jun 2018 14:42
Operator : UA/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleID :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 20 14:54:34 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062018.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 20 14:06:28 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPEST1 Signal #2 Phase: ZB-CLPEST2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.063 min
Delta R.T.: 0.000 min
Response: 4950671
Conc: 53.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.136 min
Delta R.T.: 0.010 min
Response: 12437063
Conc: 52.23 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.335 min
Delta R.T.: 0.000 min
Response: 7938652
Conc: 54.72 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.953 min
Delta R.T.: 0.020 min
Response: 8020880
Conc: 50.84 ng/ml



#3 Aroclor-1016-1


R.T.: 4.570 min
Delta R.T.: 0.001 min
Response: 1247552
Conc: 543.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 Aroclor-1016-1

R.T.: 4.490 min
Delta R.T.: 0.010 min
Response: 2037176
Conc: 501.01 ng/ml




#4 Aroclor-1016-2

R.T.: 4.995 min
Delta R.T.: 0.001 min
Response: 1423341
Conc: 537.85 ng/ml



#4 Aroclor-1016-2


R.T.: 4.960 min
Delta R.T.: 0.013 min
Response: 1544138
Conc: 512.28 ng/ml



#5 Aroclor-1016-3


R.T.: 5.041 min
Delta R.T.: 0.002 min
Response: 1003262
Conc: 552.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 Aroclor-1016-3

R.T.: 5.202 min
Delta R.T.: 0.012 min
Response: 3162459
Conc: 514.31 ng/ml




#6 Aroclor-1016-4

R.T.: 5.351 min
Delta R.T.: 0.002 min
Response: 675020
Conc: 553.86 ng/ml



#6 Aroclor-1016-4


R.T.: 5.281 min
Delta R.T.: 0.012 min
Response: 2054714
Conc: 517.40 ng/ml



#7 Aroclor-1016-5


R.T.: 5.377 min
Delta R.T.: 0.001 min
Response: 780027
Conc: 515.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 Aroclor-1016-5

R.T.: 5.545 min
Delta R.T.: 0.014 min
Response: 2284886
Conc: 550.47 ng/ml




#31 Aroclor-1260-1

R.T.: 6.193 min
Delta R.T.: 0.000 min
Response: 1825674
Conc: 512.24 ng/ml



#31 Aroclor-1260-1


R.T.: 6.514 min
Delta R.T.: 0.013 min
Response: 4153973
Conc: 505.24 ng/ml



#32 Aroclor-1260-2


R.T.: 6.629 min
Delta R.T.: 0.001 min
Response: 4701604
Conc: 513.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 Aroclor-1260-2

R.T.: 6.666 min
Delta R.T.: 0.014 min
Response: 4619124
Conc: 503.05 ng/ml




#33 Aroclor-1260-3

R.T.: 6.715 min
Delta R.T.: 0.000 min
Response: 3300701
Conc: 500.41 ng/ml



#33 Aroclor-1260-3


R.T.: 7.035 min
Delta R.T.: 0.014 min
Response: 3279348
Conc: 500.43 ng/ml



#34 Aroclor-1260-4


R.T.: 6.915 min
Delta R.T.: 0.001 min
Response: 2721447
Conc: 525.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 Aroclor-1260-4

R.T.: 7.287 min
Delta R.T.: 0.014 min
Response: 3708275
Conc: 505.33 ng/ml



#35 Aroclor-1260-5

R.T.: 7.161 min
Delta R.T.: 0.000 min
Response: 9104596
Conc: 493.59 ng/ml



#35 Aroclor-1260-5

R.T.: 7.480 min
Delta R.T.: 0.014 min
Response: 8057293
Conc: 509.52 ng/ml