

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062018\
Data File : PP026536.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Jun 2018 15:55
Operator : UA/AJ
Sample : J3532-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
OILY-WATER-DRUMS-1-2

Manual Integrations
APPROVED

ugo
6/22/2018 10:42:58 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 20 16:14:46 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062018.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 20 14:06:28 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPEST1 Signal #2 Phase: ZB-CLPEST2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.063	4.130	1846963	768076	20.018	3.226m#
2) SA Decachlor...	8.309	8.895	3932537	302200	27.107	1.915m#

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062018\
Data File : PP026536.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Jun 2018 15:55
Operator : UA/AJ
Sample : J3532-01
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
OILY-WATER-DRUMS-1-2


Manual Integrations
APPROVED

ugo
6/22/2018 10:42:58 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 20 16:14:46 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062018.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 20 14:06:28 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-CLPEST1 Signal #2 Phase: ZB-CLPEST2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.063 min
Delta R.T.: 0.000 min
Response: 1846963
Conc: 20.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
OILY-WATER-DRUMS-1-2


Manual Integrations
APPROVED

ugo
6/22/2018 10:42:58 AM




#1 Tetrachloro-m-xylene

R.T.: 4.130 min
Delta R.T.: 0.004 min
Response: 768076
Conc: 3.23 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.309 min
Delta R.T.: -0.025 min
Response: 3932537
Conc: 27.11 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.895 min
Delta R.T.: -0.038 min
Response: 302200
Conc: 1.92 ng/ml m