

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062021\
 Data File : PP036622.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Jun 2021 8:45
 Operator : DD\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 21 03:50:31 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 18 06:06:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.966	3.958	2206322	1582970	49.918	51.108
2)	SA Decachlor...	11.000	9.261	2090139	1456353	48.724	45.539
Target Compounds							
21)	L5 AR-1248-1	6.286	5.211	489416	326235	485.135	468.212
22)	L5 AR-1248-2	6.580	5.475	644479	443687	474.559	468.430
23)	L5 AR-1248-3	6.798	5.519	777786	465220	478.027	470.186
24)	L5 AR-1248-4	7.226	5.702	852897	555674	467.150	475.046
25)	L5 AR-1248-5	7.266	6.126	849612	563101	471.210	474.724



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062021\
Data File : PP036622.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Jun 2021 8:45
Operator : DD\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 03:50:31 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061721.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 18 06:06:19 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.966 min
Delta R.T.: 0.000 min
Response: 2206322
Conc: 49.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.958 min
Delta R.T.: -0.001 min
Response: 1582970
Conc: 51.11 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.000 min
Delta R.T.: 0.002 min
Response: 2090139
Conc: 48.72 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.261 min
Delta R.T.: -0.002 min
Response: 1456353
Conc: 45.54 ng/ml



#21 AR-1248-1


R.T.: 6.286 min
Delta R.T.: 0.002 min
Response: 489416
Conc: 485.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 5.211 min
Delta R.T.: -0.001 min
Response: 326235
Conc: 468.21 ng/ml




#22 AR-1248-2

R.T.: 6.580 min
Delta R.T.: 0.001 min
Response: 644479
Conc: 474.56 ng/ml



#22 AR-1248-2


R.T.: 5.475 min
Delta R.T.: -0.001 min
Response: 443687
Conc: 468.43 ng/ml



#23 AR-1248-3


R.T.: 6.798 min
Delta R.T.: 0.002 min
Response: 777786
Conc: 478.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.519 min
Delta R.T.: -0.001 min
Response: 465220
Conc: 470.19 ng/ml




#24 AR-1248-4

R.T.: 7.226 min
Delta R.T.: 0.002 min
Response: 852897
Conc: 467.15 ng/ml



#24 AR-1248-4


R.T.: 5.702 min
Delta R.T.: -0.002 min
Response: 555674
Conc: 475.05 ng/ml



#25 AR-1248-5

R.T.: 7.266 min
Delta R.T.: 0.002 min
Response: 849612
Conc: 471.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 6.126 min
Delta R.T.: 0.000 min
Response: 563101
Conc: 474.72 ng/ml