

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062021\
 Data File : PP036636.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Jun 2021 12:57
 Operator : DD\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 21 03:57:10 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 18 06:06:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.965	3.958	2145345	1506927	48.539	48.652
2)	SA Decachlor...	10.998	9.260	1760156	1246452	41.032	38.975
Target Compounds							
16)	L4 AR-1242-1	6.285	5.212	585832	408751	456.845	459.223
17)	L4 AR-1242-2	6.310	5.232	821256	559210	450.048	458.441
18)	L4 AR-1242-3	6.375	5.422	527042	309465	449.250	459.910
19)	L4 AR-1242-4	6.483	5.520	432395	289034	446.797	446.433
20)	L4 AR-1242-5	7.265	6.083	409555	366232	389.356	448.867



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062021\
Data File : PP036636.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Jun 2021 12:57
Operator : DD\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 03:57:10 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061721.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 18 06:06:19 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.965 min
Delta R.T.: 0.000 min
Response: 2145345
Conc: 48.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.958 min
Delta R.T.: 0.000 min
Response: 1506927
Conc: 48.65 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.998 min
Delta R.T.: 0.000 min
Response: 1760156
Conc: 41.03 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.260 min
Delta R.T.: -0.003 min
Response: 1246452
Conc: 38.98 ng/ml



#16 AR-1242-1


R.T.: 6.285 min
Delta R.T.: 0.000 min
Response: 585832
Conc: 456.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 5.212 min
Delta R.T.: -0.001 min
Response: 408751
Conc: 459.22 ng/ml




#17 AR-1242-2

R.T.: 6.310 min
Delta R.T.: 0.001 min
Response: 821256
Conc: 450.05 ng/ml



#17 AR-1242-2


R.T.: 5.232 min
Delta R.T.: -0.001 min
Response: 559210
Conc: 458.44 ng/ml



#18 AR-1242-3


R.T.: 6.375 min
Delta R.T.: 0.000 min
Response: 527042
Conc: 449.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.422 min
Delta R.T.: -0.001 min
Response: 309465
Conc: 459.91 ng/ml




#19 AR-1242-4

R.T.: 6.483 min
Delta R.T.: 0.000 min
Response: 432395
Conc: 446.80 ng/ml



#19 AR-1242-4


R.T.: 5.520 min
Delta R.T.: -0.001 min
Response: 289034
Conc: 446.43 ng/ml



#20 AR-1242-5

R.T.: 7.265 min
Delta R.T.: 0.000 min
Response: 409555
Conc: 389.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 6.083 min
Delta R.T.: 0.000 min
Response: 366232
Conc: 448.87 ng/ml