

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062022\
 Data File : PP048936.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jun 2022 13:43
 Operator : YP\AJ
 Sample : N3410-03 10X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 KMA786D-1-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 20 16:28:40 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 07 11:09:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.034	3.219	3317521	3581454	1.384	1.882 #
2) SA Decachlor...	10.381	8.622	3666751	2863901	1.924	2.339

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062022\
Data File : PP048936.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2022 13:43
Operator : YP\AJ
Sample : N3410-03 10X
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
KMA786D-1-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 20 16:28:40 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 07 11:09:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.034 min
Delta R.T.: 0.000 min
Response: 3317521
Conc: 1.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
KMA786D-1-1




#1 Tetrachloro-m-xylene

R.T.: 3.219 min
Delta R.T.: 0.014 min
Response: 3581454
Conc: 1.88 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.381 min
Delta R.T.: -0.005 min
Response: 3666751
Conc: 1.92 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.622 min
Delta R.T.: 0.007 min
Response: 2863901
Conc: 2.34 ng/ml