

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062023\
 Data File : PP058563.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jun 2023 22:32
 Operator : YP\AJ
 Sample : AR1242ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 21 04:32:34 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 21 04:28:27 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.460	3.671	131.8E6	87163748	50.000	50.000
2) SA Decachlor...	10.278	8.730	68449651	62803631	50.000	50.000
Target Compounds						
16) L4 AR-1242-1	5.636	4.770	28226862	20070360	500.000	500.000
17) L4 AR-1242-2	5.659	4.789	40495937	28100441	500.000	500.000
18) L4 AR-1242-3	5.722	4.967	25574571	15413905	500.000	500.000
19) L4 AR-1242-4	5.820	5.051	20252351	16639805	500.000	500.000
20) L4 AR-1242-5	6.560	5.579	18376083	19252131	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062023\
Data File : PP058563.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2023 22:32
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 04:32:34 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062023.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 21 04:28:27 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

