

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062024\
 Data File : PP065870.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jun 2024 11:38
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 21 01:45:42 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 06 05:25:59 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.829	4.066	82554173	119.9E6	43.520	42.948
2) SA Decachlor...	10.840	9.244	89350514	65636938	45.428	44.554
Target Compounds						
16) L4 AR-1242-1	6.006	5.178	23448120	28739960	428.801	399.942
17) L4 AR-1242-2	6.029	5.199	34766268	40515273	438.407	407.407
18) L4 AR-1242-3	6.093	5.380	22339883	21819180	441.952	405.645
19) L4 AR-1242-4	6.193	5.462	17961672	21869263	437.711	410.612
20) L4 AR-1242-5	6.932	5.993	18861974	24080422	432.397	405.269



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062024\
Data File : PP065870.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2024 11:38
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 01:45:42 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 06 05:25:59 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.829 min
Delta R.T.: 0.003 min
Response: 82554173
Conc: 43.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.066 min
Delta R.T.: -0.005 min
Response: 119924000
Conc: 42.95 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.840 min
Delta R.T.: -0.002 min
Response: 89350514
Conc: 45.43 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.244 min
Delta R.T.: -0.017 min
Response: 65636938
Conc: 44.55 ng/ml



#16 AR-1242-1


R.T.: 6.006 min
Delta R.T.: 0.002 min
Response: 23448120
Conc: 428.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 5.178 min
Delta R.T.: -0.008 min
Response: 28739960
Conc: 399.94 ng/ml




#17 AR-1242-2

R.T.: 6.029 min
Delta R.T.: 0.002 min
Response: 34766268
Conc: 438.41 ng/ml



#17 AR-1242-2


R.T.: 5.199 min
Delta R.T.: -0.008 min
Response: 40515273
Conc: 407.41 ng/ml



#18 AR-1242-3


R.T.: 6.093 min
Delta R.T.: 0.002 min
Response: 22339883
Conc: 441.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.380 min
Delta R.T.: -0.008 min
Response: 21819180
Conc: 405.64 ng/ml




#19 AR-1242-4

R.T.: 6.193 min
Delta R.T.: 0.002 min
Response: 17961672
Conc: 437.71 ng/ml



#19 AR-1242-4


R.T.: 5.462 min
Delta R.T.: -0.009 min
Response: 21869263
Conc: 410.61 ng/ml



#20 AR-1242-5

R.T.: 6.932 min
Delta R.T.: 0.001 min
Response: 18861974
Conc: 432.40 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.993 min
Delta R.T.: -0.009 min
Response: 24080422
Conc: 405.27 ng/ml