

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062025\
 Data File : PP073118.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jun 2025 09:33
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 20 11:38:33 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061725.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 18 08:17:15 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.496	3.786	104.1E6	84963638	49.300	47.155
2)	SA Decachlor...	10.189	8.796	88908846	63975452	52.027	47.932
Target Compounds							
16)	L4 AR-1242-1	5.647	4.867	29962372	27089565	512.727	490.584
17)	L4 AR-1242-2	5.669	4.884	45424884	39999808	524.814	503.228
18)	L4 AR-1242-3	5.731	5.061	27848237	21824548	528.353	503.973
19)	L4 AR-1242-4	5.828	5.145	23137889	21048144	526.361	507.137
20)	L4 AR-1242-5	6.558	5.667	24913740	25925902	502.702	493.014



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062025\
Data File : PP073118.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2025 09:33
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 20 11:38:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061725.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 18 08:17:15 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.496 min
Delta R.T.: 0.002 min
Response: 104062355
Conc: 49.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.786 min
Delta R.T.: -0.002 min
Response: 84963638
Conc: 47.15 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.189 min
Delta R.T.: 0.002 min
Response: 88908846
Conc: 52.03 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.796 min
Delta R.T.: -0.001 min
Response: 63975452
Conc: 47.93 ng/ml



#16 AR-1242-1


R.T.: 5.647 min
Delta R.T.: 0.002 min
Response: 29962372
Conc: 512.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.867 min
Delta R.T.: 0.000 min
Response: 27089565
Conc: 490.58 ng/ml




#17 AR-1242-2

R.T.: 5.669 min
Delta R.T.: 0.003 min
Response: 45424884
Conc: 524.81 ng/ml



#17 AR-1242-2


R.T.: 4.884 min
Delta R.T.: -0.001 min
Response: 39999808
Conc: 503.23 ng/ml



#18 AR-1242-3


R.T.: 5.731 min
Delta R.T.: 0.002 min
Response: 27848237
Conc: 528.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.061 min
Delta R.T.: 0.000 min
Response: 21824548
Conc: 503.97 ng/ml




#19 AR-1242-4

R.T.: 5.828 min
Delta R.T.: 0.003 min
Response: 23137889
Conc: 526.36 ng/ml



#19 AR-1242-4


R.T.: 5.145 min
Delta R.T.: 0.000 min
Response: 21048144
Conc: 507.14 ng/ml



#20 AR-1242-5

R.T.: 6.558 min
Delta R.T.: 0.003 min
Response: 24913740
Conc: 502.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.667 min
Delta R.T.: 0.000 min
Response: 25925902
Conc: 493.01 ng/ml