

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062025\
 Data File : PP073119.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jun 2025 09:50
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 20 11:39:11 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061725.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 18 08:17:15 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.490	3.786	100.8E6	82523491	47.763	45.801
2) SA Decachlor...	10.183	8.794	88605870	65851503	51.850	49.337
Target Compounds						
21) L5 AR-1248-1	5.642	4.866	22286448	20881442	468.513	479.038
22) L5 AR-1248-2	5.914	5.102	30021073	28427693	512.734	485.151
23) L5 AR-1248-3	6.116	5.144	34011523	29703821	521.255	490.614
24) L5 AR-1248-4	6.515	5.316	41293878	34795209	485.442	496.915
25) L5 AR-1248-5	6.553	5.707	40020026	34802379	478.176	489.954



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062025\
Data File : PP073119.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2025 09:50
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 20 11:39:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061725.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 18 08:17:15 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.490 min
Delta R.T.: -0.004 min
Response: 100818498
Conc: 47.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.786 min
Delta R.T.: -0.001 min
Response: 82523491
Conc: 45.80 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.183 min
Delta R.T.: -0.004 min
Response: 88605870
Conc: 51.85 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.794 min
Delta R.T.: -0.003 min
Response: 65851503
Conc: 49.34 ng/ml



#21 AR-1248-1


R.T.: 5.642 min
Delta R.T.: -0.003 min
Response: 22286448
Conc: 468.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.866 min
Delta R.T.: -0.002 min
Response: 20881442
Conc: 479.04 ng/ml




#22 AR-1248-2

R.T.: 5.914 min
Delta R.T.: -0.003 min
Response: 30021073
Conc: 512.73 ng/ml



#22 AR-1248-2


R.T.: 5.102 min
Delta R.T.: -0.001 min
Response: 28427693
Conc: 485.15 ng/ml



#23 AR-1248-3


R.T.: 6.116 min
Delta R.T.: -0.003 min
Response: 34011523
Conc: 521.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.144 min
Delta R.T.: -0.001 min
Response: 29703821
Conc: 490.61 ng/ml




#24 AR-1248-4

R.T.: 6.515 min
Delta R.T.: -0.002 min
Response: 41293878
Conc: 485.44 ng/ml



#24 AR-1248-4


R.T.: 5.316 min
Delta R.T.: -0.002 min
Response: 34795209
Conc: 496.91 ng/ml



#25 AR-1248-5

R.T.: 6.553 min
Delta R.T.: -0.002 min
Response: 40020026
Conc: 478.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.707 min
Delta R.T.: -0.002 min
Response: 34802379
Conc: 489.95 ng/ml