

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062122\
 Data File : PP048964.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jun 2022 16:09
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/22/2022
 Supervised By :Ankita Jodhani 06/22/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 21 18:22:45 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 07 11:09:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.033	3.198	120.1E6	112.6E6	50.140m	59.174
2)	SA Decachlor...	10.379	8.601	94124445	61827160	49.383	50.489
Target Compounds							
3)	L1 AR-1016-1	5.321	4.352	39202814	40571073	455.655	585.618m#
4)	L1 AR-1016-2	5.345	4.372	56887928	54290729	467.252	572.397
5)	L1 AR-1016-3	5.412	4.558	34630251	29821986	456.896	573.370 #
6)	L1 AR-1016-4	5.521	4.606	28483473	23432636	452.802	582.148 #
7)	L1 AR-1016-5	5.845	4.832	27883867	28952891	451.667	567.721 #
31)	L7 AR-1260-1	7.090	5.948	46997195	46281128	451.047	548.477
32)	L7 AR-1260-2	7.377	6.154	55354096	54034747	454.592	527.532
33)	L7 AR-1260-3	7.774	6.319	40004031	49565717	472.526	511.451
34)	L7 AR-1260-4	8.026	6.833	48172493	37238086	465.605	526.128
35)	L7 AR-1260-5	8.389	7.099	91884171	83201911	469.378	518.095

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062122\
Data File : PP048964.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jun 2022 16:09
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/22/2022
Supervised By :Ankita Jodhani 06/22/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 21 18:22:45 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 07 11:09:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

