

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062122\
 Data File : PP048976.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jun 2022 19:41
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/22/2022
 Supervised By :Ankita Jodhani 06/22/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 22 00:46:41 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 07 11:09:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.033	3.198	84053096	111.7E6	35.078m	58.666 #
2) SA Decachlor...	10.379	8.603	78240164	49253642	41.049	40.221
Target Compounds						
3) L1 AR-1016-1	5.320	4.352	27600610	33462411	320.803	483.009m#
4) L1 AR-1016-2	5.345	4.372	40313832	45470518	331.120	479.404 #
5) L1 AR-1016-3	5.412	4.558	24515388	24654372	323.445	474.015 #
6) L1 AR-1016-4	5.521	4.606	20250491	19315241	321.922	479.857 #
7) L1 AR-1016-5	5.845	4.832	19659621	23684428	318.450	464.415 #
31) L7 AR-1260-1	7.090	5.948	35560463	36121787	341.285	428.079 #
32) L7 AR-1260-2	7.378	6.154	43878145	41652090	360.347	406.643
33) L7 AR-1260-3	7.776	6.319	35006959	38713805	413.501	399.474
34) L7 AR-1260-4	8.027	6.834	43412187	28589625	419.595	403.936
35) L7 AR-1260-5	8.390	7.100	79338109	64376787	405.288	400.872

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062122\
Data File : PP048976.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jun 2022 19:41
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/22/2022
Supervised By :Ankita Jodhani 06/22/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 22 00:46:41 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 07 11:09:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

