

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062122\
 Data File : PP048974.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jun 2022 18:42
 Operator : YP\AJ
 Sample : N3422-01
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 WC-14A-1A-12

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/22/2022
 Supervised By :Ankita Jodhani 06/22/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 22 00:44:50 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 07 11:09:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.039	3.205	47903320	42968894	19.991	22.575
2) SA Decachlor...	10.405	8.633	43834974	23559061	22.998m	19.239m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062122\
Data File : PP048974.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jun 2022 18:42
Operator : YP\AJ
Sample : N3422-01
Misc :
ALS Vial : 22 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
WC-14A-1A-12


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/22/2022
Supervised By :Ankita Jodhani 06/22/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 22 00:44:50 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060722.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 07 11:09:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm





#1 Tetrachloro-m-xylene

R.T.: 4.039 min
Delta R.T.: 0.004 min
Response: 47903320
Conc: 19.99 ng/ml

Instrument :


ECD_P

ClientSampleId :

WC-14A-1A-12


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/22/2022
Supervised By :Ankita Jodhani 06/22/2022




#1 Tetrachloro-m-xylene

R.T.: 3.205 min
Delta R.T.: 0.000 min
Response: 42968894
Conc: 22.58 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.405 min
Delta R.T.: 0.018 min
Response: 43834974
Conc: 23.00 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.633 min
Delta R.T.: 0.018 min
Response: 23559061
Conc: 19.24 ng/ml m