

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD P\methods\
 Method File : PP062218.M
 Title : 8080.M
 Last Update : Fri Jun 22 02:12:49 2018
 Response Via : Initial Calibration

Calibration Files

750 =PP026608.D 1000 =PP026609.D 1500 =PP026610.D
 500 =PP026607.D 200 =PP026606.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	5.518	5.295	5.082	5.453	5.852	5.440	E5	5.24
2) T	3,5-DICHLOROBENZOI...	5.214	5.007	4.684	5.181	5.727	5.162	E5	7.34
3) T	4-Nitrophenol	2.780	2.667	2.491	2.755	3.058	2.750	E5	7.50
4) S	2,4-DCAA	3.456	3.327	3.115	3.447	3.850	3.439	E5	7.78
5) T	DICAMBA	1.449	1.404	1.326	1.417	1.496	1.419	E6	4.41
6) T	MCPP	9.218	9.062	8.731	8.729	8.639	8.876	E5	2.82
7) T	MCPA	1.467	1.414	1.348	1.400	1.492	1.424	E6	3.99
8) T	DICHLORPROP	3.691	3.522	3.306	3.704	4.216	3.688	E5	9.12
9) T	2,4-D	4.616	4.440	4.180	4.620	5.243	4.620	E5	8.48
10) T	Pentachlorophenol	6.024	5.762	5.206	6.011	6.430	5.887	E6	7.63
11) T	2,4,5-TP (SILVEX)	2.381	2.295	2.147	2.329	2.430	2.316	E6	4.66
12) T	2,4,5-T	2.080	2.028	1.919	2.026	2.052	2.021	E6	3.02
13) T	2,4-DB	2.835	2.714	2.546	2.809	3.135	2.808	E5	7.66
14) T	DINOSEB	1.432	1.387	1.338	1.349	1.352	1.371	E6	2.80
15) T	Picloram	3.313	3.281	3.141	3.240	3.183	3.232	E6	2.17
16) T	DCPA	2.938	2.831	2.627	2.871	2.987	2.851	E6	4.86

Signal #2 Calibration Files

750 =PP026608.D 1000 =PP026609.D 1500 =PP026610.D
 500 =PP026607.D 200 =PP026606.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	5.828	5.513	5.431	5.784	6.284	5.768	E5	5.80
2) T	3,5-DICHLOROBENZOI...	5.247	4.829	4.672	5.236	5.981	5.193	E5	9.77
3) T	4-Nitrophenol	2.941	2.735	2.564	3.003	3.486	2.946	E5	11.82
4) S	2,4-DCAA	3.727	3.431	3.300	3.751	4.329	3.708	E5	10.71
5) T	DICAMBA	1.445	1.335	1.287	1.429	1.562	1.412	E6	7.54
6) T	MCPP	0.953	0.879	0.850	0.952	1.027	0.932	E6	7.46
7) T	MCPA	1.532	1.404	1.331	1.588	1.942	1.559	E6	15.19
8) T	DICHLORPROP	3.451	3.177	3.037	3.525	4.107	3.459	E5	11.94
9) T	2,4-D	4.332	4.025	3.890	4.391	4.981	4.324	E5	9.78
10) T	Pentachlorophenol	5.420	4.945	4.531	5.579	6.219	5.339	E6	12.02
11) T	2,4,5-TP (SILVEX)	2.038	1.900	1.809	2.049	2.228	2.005	E6	7.97
12) T	2,4,5-T	1.617	1.527	1.476	1.594	1.702	1.583	E6	5.47
13) T	2,4-DB	2.186	2.042	1.947	2.205	2.504	2.177	E5	9.73
14) T	DINOSEB	1.074	1.014	0.999	1.036	1.098	1.044	E6	3.98
15) T	Picloram	1.943	1.921	1.894	1.854	1.788	1.880	E6	3.26
16) T	DCPA	2.284	2.159	2.035	2.289	2.489	2.251	E6	7.52

(#) = Out of Range