

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062221\
 Data File : PP036694.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Jun 2021 18:54
 Operator : DD\AJ
 Sample : M2789-07
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 SHORT-PILLARS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 23 06:38:08 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 18 06:06:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.963	3.956	722901	501847	16.356	16.203
2)	SA Decachlor...	10.998	9.257	737614	463127	17.195	14.482
Target Compounds							
8)	L2 AR-1221-1	5.214	0.000	17236	0	30.701	N.D. #
9)	L2 AR-1221-2	5.316	0.000	11979	0	27.798	N.D. #
24)	L5 AR-1248-4	7.199f	0.000	19731	0	10.807	N.D. #
26)	L6 AR-1254-1	7.199	0.000	19731	0	11.219	N.D. #
27)	L6 AR-1254-2	7.426	0.000	10545	0	3.897	N.D. #
28)	L6 AR-1254-3	7.815	0.000	21109	0	7.127	N.D. #
32)	L7 AR-1260-2	0.000	6.997	0	6251	N.D.	3.223 #
35)	L7 AR-1260-5	0.000	7.892	0	12674	N.D.	3.449 #
37)	L8 AR-1262-2	0.000	7.892	0	12674	N.D.	3.233 #
39)	L8 AR-1262-4	9.544f	0.000	13287	0	4.626	N.D. #
42)	L9 AR-1268-2	9.544f	0.000	13287	0	2.331	N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062221\
Data File : PP036694.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Jun 2021 18:54
Operator : DD\AJ
Sample : M2789-07
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
SHORT-PILLARS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 23 06:38:08 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061721.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 18 06:06:19 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.963 min
Delta R.T.: -0.002 min
Response: 722901
Conc: 16.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
SHORT-PILLARS




#1 Tetrachloro-m-xylene

R.T.: 3.956 min
Delta R.T.: -0.003 min
Response: 501847
Conc: 16.20 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.998 min
Delta R.T.: 0.000 min
Response: 737614
Conc: 17.19 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.257 min
Delta R.T.: -0.006 min
Response: 463127
Conc: 14.48 ng/ml



#8 AR-1221-1


R.T.: 5.214 min
Delta R.T.: -0.002 min
Response: 17236
Conc: 30.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
SHORT-PILLARS




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.217 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 5.316 min
Delta R.T.: 0.000 min
Response: 11979
Conc: 27.80 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T. : 4.316 min
Response: 0
Conc: N.D.



#24 AR-1248-4


R.T.: 7.199 min
Delta R.T.: -0.025 min
Response: 19731
Conc: 10.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
SHORT-PILLARS




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T. : 5.704 min
Response: 0
Conc: N.D.




#26 AR-1254-1

R.T.: 7.199 min
Delta R.T.: 0.005 min
Response: 19731
Conc: 11.22 ng/ml



#26 AR-1254-1


R.T.: 0.000 min
Exp R.T. : 6.084 min
Response: 0
Conc: N.D.



#27 AR-1254-2


R.T.: 7.426 min
Delta R.T.: 0.002 min
Response: 10545
Conc: 3.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
SHORT-PILLARS




#27 AR-1254-2

R.T.: 0.000 min
Exp R.T. : 6.245 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 7.815 min
Delta R.T.: 0.010 min
Response: 21109
Conc: 7.13 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 6.665 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T. : 8.238 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
SHORT-PILLARS




#32 AR-1260-2

R.T.: 6.997 min
Delta R.T.: -0.004 min
Response: 6251
Conc: 3.22 ng/ml




#35 AR-1260-5

R.T.: 0.000 min
Exp R.T. : 9.163 min
Response: 0
Conc: N.D.



#35 AR-1260-5


R.T.: 7.892 min
Delta R.T.: 0.005 min
Response: 12674
Conc: 3.45 ng/ml



#37 AR-1262-2


R.T.: 0.000 min
Exp R.T. : 9.162 min
Response: 0
Conc: N.D.

Instrument :
ECD_P
ClientSampleId :
SHORT-PILLARS




#37 AR-1262-2

R.T.: 7.892 min
Delta R.T.: 0.006 min
Response: 12674
Conc: 3.23 ng/ml




#39 AR-1262-4

R.T.: 9.544 min
Delta R.T.: -0.023 min
Response: 13287
Conc: 4.63 ng/ml



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T. : 8.238 min
Response: 0
Conc: N.D.



#42 AR-1268-2

R.T.: 9.544 min
Delta R.T.: -0.027 min
Response: 13287
Conc: 2.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
SHORT-PILLARS



#42 AR-1268-2

R.T.: 0.000 min
Exp R.T. : 8.240 min
Response: 0
Conc: N.D.