

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062221\
 Data File : PP036677.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Jun 2021 13:35
 Operator : DD\AJ
 Sample : AR1268CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 22 14:37:03 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 18 06:06:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.966	3.957	2281369	1643610	51.616	53.065
2) SA Decachlor...	11.002	9.259	3633012	2375386	84.691	74.276
Target Compounds						
41) L9 AR-1268-1	9.477	8.173	2677380	1981597	446.266	425.547
42) L9 AR-1268-2	9.574	8.238	2564345	1827892	449.857	423.753
43) L9 AR-1268-3	9.801	8.447	2198223	1552546	457.230m	425.226
44) L9 AR-1268-4	10.245	8.735	1026014	665665	435.583	418.467
45) L9 AR-1268-5	10.663	9.015	7334479	5023954	476.937	424.943

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Instrument :
 ECD_P
 Client Sample ID :
 AR1268CCC500

Manual Integrations
 APPROVED

Ankita
 6/23/2021 2:17:11 PM

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062221\
Data File : PP036677.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Jun 2021 13:35
Operator : DD\AJ
Sample : AR1268CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 22 14:37:03 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061721.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 18 06:06:19 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Instrument :

ECD_P

ClientSampled :



AR1268CCC500


Manual Integrations
APPROVED

Ankita

6/23/2021 2:17:11 PM

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.966 min
Delta R.T.: 0.000 min
Response: 2281369
Conc: 51.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268CCC500


Manual Integrations
APPROVED

Ankita
6/23/2021 2:17:11 PM




#1 Tetrachloro-m-xylene

R.T.: 3.957 min
Delta R.T.: -0.002 min
Response: 1643610
Conc: 53.07 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.002 min
Delta R.T.: 0.005 min
Response: 3633012
Conc: 84.69 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.259 min
Delta R.T.: -0.004 min
Response: 2375386
Conc: 74.28 ng/ml




#41 AR-1268-1

R.T.: 9.477 min
Delta R.T.: 0.004 min
Response: 2677380
Conc: 446.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268CCC500


Manual Integrations
APPROVED

Ankita
6/23/2021 2:17:11 PM




#41 AR-1268-1

R.T.: 8.173 min
Delta R.T.: -0.002 min
Response: 1981597
Conc: 425.55 ng/ml




#42 AR-1268-2

R.T.: 9.574 min
Delta R.T.: 0.003 min
Response: 2564345
Conc: 449.86 ng/ml



#42 AR-1268-2

R.T.: 8.238 min
Delta R.T.: -0.002 min
Response: 1827892
Conc: 423.75 ng/ml




#43 AR-1268-3

R.T.: 9.801 min
Delta R.T.: 0.003 min
Response: 2198223
Conc: 457.23 ng/ml m

Instrument :
ECD_P
ClientSampleId :
AR1268CCC500


Manual Integrations
APPROVED

Ankita
6/23/2021 2:17:11 PM




#43 AR-1268-3

R.T.: 8.447 min
Delta R.T.: -0.001 min
Response: 1552546
Conc: 425.23 ng/ml




#44 AR-1268-4

R.T.: 10.245 min
Delta R.T.: 0.005 min
Response: 1026014
Conc: 435.58 ng/ml



#44 AR-1268-4

R.T.: 8.735 min
Delta R.T.: 0.000 min
Response: 665665
Conc: 418.47 ng/ml




#45 AR-1268-5

R.T.: 10.663 min
Delta R.T.: 0.005 min
Response: 7334479
Conc: 476.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268CCC500

Manual Integrations
APPROVED

Ankita
6/23/2021 2:17:11 PM



#45 AR-1268-5

R.T.: 9.015 min
Delta R.T.: -0.001 min
Response: 5023954
Conc: 424.94 ng/ml