

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062223\
 Data File : PP058620.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Jun 2023 10:46
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 22 17:22:54 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 21 05:48:37 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.459	3.669	109.5E6	87762138	43.857	53.807
2)	SA Decachlor...	10.278	8.724	66473927	70349578	50.259	57.757
Target Compounds							
3)	L1 AR-1016-1	5.635	4.767	31828469	28139250	445.881	559.096 #
4)	L1 AR-1016-2	5.658	4.786	46373163	39951397	449.725	564.371 #
5)	L1 AR-1016-3	5.721	4.964	28841441	21392025	449.624	567.933 #
6)	L1 AR-1016-4	5.819	5.006	23047388	18271859	450.074	549.382
7)	L1 AR-1016-5	6.115	5.220	22957970	22787823	434.691	560.202 #
8)	L2 AR-1221-1	4.663	3.885	3699998	2916068	126.592	142.199
9)	L2 AR-1221-2	4.756	3.972	5378267	4226353	245.967	268.679
10)	L2 AR-1221-3	4.832	4.050	18406328	16541049	286.626	347.389
11)	L3 AR-1232-1	4.832	4.050	18406328	16541049	396.007	482.138
12)	L3 AR-1232-2	5.366	4.786	24760528	39951397	923.527	1234.009 #
13)	L3 AR-1232-3	5.658	4.964	46373163	21392025	1011.856	1230.663
14)	L3 AR-1232-4	5.819	5.048	23047388	22124284	992.732	1315.896 #
15)	L3 AR-1232-5	5.910	5.220	20683977	22787823	1058.105	1232.385
16)	L4 AR-1242-1	5.635	4.767	31828469	28139250	568.944	709.196
17)	L4 AR-1242-2	5.658	4.786	46373163	39951397	577.575	718.905
18)	L4 AR-1242-3	5.721	4.964	28841441	21392025	569.764	727.507 #
19)	L4 AR-1242-4	5.819	5.048	23047388	22124284	576.782	684.167
20)	L4 AR-1242-5	6.560	5.576	3040945	18104142	83.671	477.344 #
21)	L5 AR-1248-1	5.635	4.767	31828469	28139250	749.527	921.161
22)	L5 AR-1248-2	5.910	5.006	20683977	18271859	320.837	399.335
23)	L5 AR-1248-3	6.115	5.048	22957970	22124284	325.874	466.380 #
24)	L5 AR-1248-4	6.519	5.220	3142019	22787823	46.018	416.282 #
25)	L5 AR-1248-5	6.560	5.616	3040945	2524288	47.440	51.433
26)	L6 AR-1254-1	6.495	5.576	17293576	18104142	225.561	234.658
27)	L6 AR-1254-2	6.713	5.724	18092309	17462969	163.354	255.071 #
28)	L6 AR-1254-3	7.082	6.146	9423544	30494224	87.857	290.553 #
29)	L6 AR-1254-4	7.368	6.359	5920808	3000242	81.448	50.212 #
30)	L6 AR-1254-5	7.787	6.779	52976183	57148544	655.436	618.000
31)	L7 AR-1260-1	7.246	6.261	39141509	42829461	448.202	554.191
32)	L7 AR-1260-2	7.504	6.450	44340991	50168646	459.088	567.304
33)	L7 AR-1260-3	7.865	6.604	28507144	47266657	466.472	560.575
34)	L7 AR-1260-4	8.090	7.079	35295047	34119143	476.131	565.938
35)	L7 AR-1260-5	8.415	7.321	63723728	73254708	489.177	572.621
36)	L8 AR-1262-1	7.865	6.872	28507144	18946158	277.115	445.993 #
37)	L8 AR-1262-2	8.415	7.079	63723728	34119143	378.840	377.670
38)	L8 AR-1262-3	8.747	7.607	41259056	15470675	361.592	226.892 #
39)	L8 AR-1262-4	8.825	7.670	9807893	54020492	113.892	442.111 #
40)	L8 AR-1262-5	9.497	8.168	14577591	15816364	265.765	299.140
41)	L9 AR-1268-1	8.747f	7.607	41259056	15470675	210.220	83.808 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062223\
 Data File : PP058620.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Jun 2023 10:46
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 22 17:22:54 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062023.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 21 05:48:37 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.825	7.670	9807893	54020492	56.654	327.458 #
43) L9	AR-1268-3	9.063	7.882	1169928	1302599	7.535	8.998
44) L9	AR-1268-4	9.497	8.168	14577591	15816364	254.847	290.670
45) L9	AR-1268-5	9.927	8.463	5154448	5630317	11.061	13.520



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062223\
Data File : PP058620.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jun 2023 10:46
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 22 17:22:54 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062023.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 21 05:48:37 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.459 min
Delta R.T.: 0.000 min
Response: 109519180
Conc: 43.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.669 min
Delta R.T.: -0.002 min
Response: 87762138
Conc: 53.81 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.278 min
Delta R.T.: 0.003 min
Response: 66473927
Conc: 50.26 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.724 min
Delta R.T.: -0.005 min
Response: 70349578
Conc: 57.76 ng/ml



#3 AR-1016-1


R.T.: 5.635 min
Delta R.T.: 0.000 min
Response: 31828469
Conc: 445.88 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.767 min
Delta R.T.: -0.003 min
Response: 28139250
Conc: 559.10 ng/ml




#4 AR-1016-2

R.T.: 5.658 min
Delta R.T.: 0.000 min
Response: 46373163
Conc: 449.73 ng/ml



#4 AR-1016-2


R.T.: 4.786 min
Delta R.T.: -0.003 min
Response: 39951397
Conc: 564.37 ng/ml



#5 AR-1016-3


R.T.: 5.721 min
Delta R.T.: 0.000 min
Response: 28841441
Conc: 449.62 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.964 min
Delta R.T.: -0.003 min
Response: 21392025
Conc: 567.93 ng/ml




#6 AR-1016-4

R.T.: 5.819 min
Delta R.T.: 0.000 min
Response: 23047388
Conc: 450.07 ng/ml



#6 AR-1016-4


R.T.: 5.006 min
Delta R.T.: -0.003 min
Response: 18271859
Conc: 549.38 ng/ml



#7 AR-1016-5


R.T.: 6.115 min
Delta R.T.: 0.000 min
Response: 22957970
Conc: 434.69 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.220 min
Delta R.T.: -0.003 min
Response: 22787823
Conc: 560.20 ng/ml




#8 AR-1221-1

R.T.: 4.663 min
Delta R.T.: -0.005 min
Response: 3699998
Conc: 126.59 ng/ml



#8 AR-1221-1


R.T.: 3.885 min
Delta R.T.: -0.004 min
Response: 2916068
Conc: 142.20 ng/ml



#9 AR-1221-2


R.T.: 4.756 min
Delta R.T.: 0.000 min
Response: 5378267
Conc: 245.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.972 min
Delta R.T.: -0.003 min
Response: 4226353
Conc: 268.68 ng/ml




#10 AR-1221-3

R.T.: 4.832 min
Delta R.T.: 0.000 min
Response: 18406328
Conc: 286.63 ng/ml



#10 AR-1221-3


R.T.: 4.050 min
Delta R.T.: -0.003 min
Response: 16541049
Conc: 347.39 ng/ml



#11 AR-1232-1


R.T.: 4.832 min
Delta R.T.: 0.000 min
Response: 18406328
Conc: 396.01 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.050 min
Delta R.T.: -0.003 min
Response: 16541049
Conc: 482.14 ng/ml




#12 AR-1232-2

R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 24760528
Conc: 923.53 ng/ml



#12 AR-1232-2


R.T.: 4.786 min
Delta R.T.: -0.003 min
Response: 39951397
Conc: 1234.01 ng/ml



#13 AR-1232-3


R.T.: 5.658 min
Delta R.T.: 0.000 min
Response: 46373163
Conc: 1011.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.964 min
Delta R.T.: -0.003 min
Response: 21392025
Conc: 1230.66 ng/ml




#14 AR-1232-4

R.T.: 5.819 min
Delta R.T.: 0.000 min
Response: 23047388
Conc: 992.73 ng/ml



#14 AR-1232-4


R.T.: 5.048 min
Delta R.T.: -0.004 min
Response: 22124284
Conc: 1315.90 ng/ml



#15 AR-1232-5


R.T.: 5.910 min
Delta R.T.: -0.001 min
Response: 20683977
Conc: 1058.11 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.220 min
Delta R.T.: -0.004 min
Response: 22787823
Conc: 1232.38 ng/ml




#16 AR-1242-1

R.T.: 5.635 min
Delta R.T.: 0.000 min
Response: 31828469
Conc: 568.94 ng/ml



#16 AR-1242-1


R.T.: 4.767 min
Delta R.T.: -0.003 min
Response: 28139250
Conc: 709.20 ng/ml



#17 AR-1242-2


R.T.: 5.658 min
Delta R.T.: 0.000 min
Response: 46373163
Conc: 577.58 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.786 min
Delta R.T.: -0.002 min
Response: 39951397
Conc: 718.91 ng/ml




#18 AR-1242-3

R.T.: 5.721 min
Delta R.T.: 0.000 min
Response: 28841441
Conc: 569.76 ng/ml



#18 AR-1242-3


R.T.: 4.964 min
Delta R.T.: -0.003 min
Response: 21392025
Conc: 727.51 ng/ml



#19 AR-1242-4


R.T.: 5.819 min
Delta R.T.: 0.000 min
Response: 23047388
Conc: 576.78 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.048 min
Delta R.T.: -0.003 min
Response: 22124284
Conc: 684.17 ng/ml




#20 AR-1242-5

R.T.: 6.560 min
Delta R.T.: 0.000 min
Response: 3040945
Conc: 83.67 ng/ml



#20 AR-1242-5


R.T.: 5.576 min
Delta R.T.: -0.002 min
Response: 18104142
Conc: 477.34 ng/ml



#21 AR-1248-1


R.T.: 5.635 min
Delta R.T.: 0.000 min
Response: 31828469
Conc: 749.53 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.767 min
Delta R.T.: -0.003 min
Response: 28139250
Conc: 921.16 ng/ml




#22 AR-1248-2

R.T.: 5.910 min
Delta R.T.: 0.000 min
Response: 20683977
Conc: 320.84 ng/ml



#22 AR-1248-2


R.T.: 5.006 min
Delta R.T.: -0.003 min
Response: 18271859
Conc: 399.34 ng/ml



#23 AR-1248-3


R.T.: 6.115 min
Delta R.T.: 0.000 min
Response: 22957970
Conc: 325.87 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.048 min
Delta R.T.: -0.003 min
Response: 22124284
Conc: 466.38 ng/ml




#24 AR-1248-4

R.T.: 6.519 min
Delta R.T.: -0.002 min
Response: 3142019
Conc: 46.02 ng/ml



#24 AR-1248-4


R.T.: 5.220 min
Delta R.T.: -0.004 min
Response: 22787823
Conc: 416.28 ng/ml



#25 AR-1248-5


R.T.: 6.560 min
Delta R.T.: 0.000 min
Response: 3040945
Conc: 47.44 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.616 min
Delta R.T.: -0.004 min
Response: 2524288
Conc: 51.43 ng/ml




#26 AR-1254-1

R.T.: 6.495 min
Delta R.T.: -0.001 min
Response: 17293576
Conc: 225.56 ng/ml



#26 AR-1254-1


R.T.: 5.576 min
Delta R.T.: -0.003 min
Response: 18104142
Conc: 234.66 ng/ml



#27 AR-1254-2


R.T.: 6.713 min
Delta R.T.: -0.002 min
Response: 18092309
Conc: 163.35 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.724 min
Delta R.T.: -0.004 min
Response: 17462969
Conc: 255.07 ng/ml




#28 AR-1254-3

R.T.: 7.082 min
Delta R.T.: 0.000 min
Response: 9423544
Conc: 87.86 ng/ml



#28 AR-1254-3


R.T.: 6.146 min
Delta R.T.: 0.012 min
Response: 30494224
Conc: 290.55 ng/ml



#29 AR-1254-4


R.T.: 7.368 min
Delta R.T.: 0.000 min
Response: 5920808
Conc: 81.45 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.359 min
Delta R.T.: -0.004 min
Response: 3000242
Conc: 50.21 ng/ml




#30 AR-1254-5

R.T.: 7.787 min
Delta R.T.: -0.001 min
Response: 52976183
Conc: 655.44 ng/ml



#30 AR-1254-5


R.T.: 6.779 min
Delta R.T.: -0.004 min
Response: 57148544
Conc: 618.00 ng/ml



#31 AR-1260-1


R.T.: 7.246 min
Delta R.T.: 0.000 min
Response: 39141509
Conc: 448.20 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.261 min
Delta R.T.: -0.004 min
Response: 42829461
Conc: 554.19 ng/ml




#32 AR-1260-2

R.T.: 7.504 min
Delta R.T.: 0.000 min
Response: 44340991
Conc: 459.09 ng/ml



#32 AR-1260-2


R.T.: 6.450 min
Delta R.T.: -0.004 min
Response: 50168646
Conc: 567.30 ng/ml



#33 AR-1260-3


R.T.: 7.865 min
Delta R.T.: 0.001 min
Response: 28507144
Conc: 466.47 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.604 min
Delta R.T.: -0.005 min
Response: 47266657
Conc: 560.58 ng/ml




#34 AR-1260-4

R.T.: 8.090 min
Delta R.T.: 0.000 min
Response: 35295047
Conc: 476.13 ng/ml



#34 AR-1260-4


R.T.: 7.079 min
Delta R.T.: -0.004 min
Response: 34119143
Conc: 565.94 ng/ml



#35 AR-1260-5


R.T.: 8.415 min
Delta R.T.: 0.001 min
Response: 63723728
Conc: 489.18 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.321 min
Delta R.T.: -0.005 min
Response: 73254708
Conc: 572.62 ng/ml




#36 AR-1262-1

R.T.: 7.865 min
Delta R.T.: 0.000 min
Response: 28507144
Conc: 277.11 ng/ml



#36 AR-1262-1


R.T.: 6.872 min
Delta R.T.: -0.003 min
Response: 18946158
Conc: 445.99 ng/ml



#37 AR-1262-2


R.T.: 8.415 min
Delta R.T.: 0.000 min
Response: 63723728
Conc: 378.84 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.079 min
Delta R.T.: -0.004 min
Response: 34119143
Conc: 377.67 ng/ml




#38 AR-1262-3

R.T.: 8.747 min
Delta R.T.: 0.013 min
Response: 41259056
Conc: 361.59 ng/ml



#38 AR-1262-3


R.T.: 7.607 min
Delta R.T.: -0.003 min
Response: 15470675
Conc: 226.89 ng/ml



#39 AR-1262-4


R.T.: 8.825 min
Delta R.T.: 0.001 min
Response: 9807893
Conc: 113.89 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.670 min
Delta R.T.: -0.005 min
Response: 54020492
Conc: 442.11 ng/ml




#40 AR-1262-5

R.T.: 9.497 min
Delta R.T.: 0.000 min
Response: 14577591
Conc: 265.76 ng/ml



#40 AR-1262-5


R.T.: 8.168 min
Delta R.T.: -0.003 min
Response: 15816364
Conc: 299.14 ng/ml



#41 AR-1268-1


R.T.: 8.747 min
Delta R.T.: 0.015 min
Response: 41259056
Conc: 210.22 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.607 min
Delta R.T.: -0.004 min
Response: 15470675
Conc: 83.81 ng/ml




#42 AR-1268-2

R.T.: 8.825 min
Delta R.T.: -0.002 min
Response: 9807893
Conc: 56.65 ng/ml



#42 AR-1268-2


R.T.: 7.670 min
Delta R.T.: -0.005 min
Response: 54020492
Conc: 327.46 ng/ml



#43 AR-1268-3


R.T.: 9.063 min
Delta R.T.: 0.002 min
Response: 1169928
Conc: 7.54 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.882 min
Delta R.T.: -0.003 min
Response: 1302599
Conc: 9.00 ng/ml




#44 AR-1268-4

R.T.: 9.497 min
Delta R.T.: 0.000 min
Response: 14577591
Conc: 254.85 ng/ml



#44 AR-1268-4


R.T.: 8.168 min
Delta R.T.: -0.004 min
Response: 15816364
Conc: 290.67 ng/ml



#45 AR-1268-5

R.T.: 9.927 min
Delta R.T.: 0.004 min
Response: 5154448
Conc: 11.06 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.463 min
Delta R.T.: -0.004 min
Response: 5630317
Conc: 13.52 ng/ml