

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062424\
 Data File : PP065975.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jun 2024 17:47
 Operator : YP\AJ
 Sample : P3023-01
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 S01-01B

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/25/2024
 Supervised By :Ankita Jodhani 06/25/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 25 02:26:09 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 06 05:25:59 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.828	4.066	40567586	51571339	21.386	18.469
2) SA Decachlor...	10.846	9.248	43434203	35063700	22.083	23.801
Target Compounds						
26) L6 AR-1254-1	6.867	5.996	188.3E6	219.2E6	2403.264m	1778.768 #
27) L6 AR-1254-2	7.084	6.143	473.8E6	414.3E6	4010.868	3934.604
28) L6 AR-1254-3	7.454	6.555	858.3E6	948.7E6	7033.029	5995.403
29) L6 AR-1254-4	7.739	6.783	610.9E6	515.2E6	6870.129	5580.063
30) L6 AR-1254-5	8.161	7.205	1816.8E6	1547.8E6	18546.705	12093.235 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062424\
Data File : PP065975.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jun 2024 17:47
Operator : YP\AJ
Sample : P3023-01
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
S01-01B

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/25/2024
Supervised By :Ankita Jodhani 06/25/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 25 02:26:09 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 06 05:25:59 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

