

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062424\
 Data File : PP065952.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jun 2024 10:50
 Operator : YP\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 24 14:03:28 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 06 05:25:59 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.829	4.066	94427956	131.6E6	49.779	47.121
2)	SA Decachlor...	10.848	9.249	104.6E6	74946266	53.158	50.873
Target Compounds							
21)	L5 AR-1248-1	6.007	5.179	21113460	25035015	519.150	469.048
22)	L5 AR-1248-2	6.281	5.419	31620722	33083239	507.367	449.472
23)	L5 AR-1248-3	6.489	5.463	34347782	34441449	510.065	450.699
24)	L5 AR-1248-4	6.893	5.639	37125233	39596843	521.457	450.401
25)	L5 AR-1248-5	6.933	6.038	36174651	35770657	521.841	448.382



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062424\
Data File : PP065952.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jun 2024 10:50
Operator : YP\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 24 14:03:28 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 06 05:25:59 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.829 min
Delta R.T.: 0.003 min
Response: 94427956
Conc: 49.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.066 min
Delta R.T.: -0.005 min
Response: 131577835
Conc: 47.12 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.848 min
Delta R.T.: 0.005 min
Response: 104554900
Conc: 53.16 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.249 min
Delta R.T.: -0.012 min
Response: 74946266
Conc: 50.87 ng/ml



#21 AR-1248-1


R.T.: 6.007 min
Delta R.T.: 0.003 min
Response: 21113460
Conc: 519.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 5.179 min
Delta R.T.: -0.006 min
Response: 25035015
Conc: 469.05 ng/ml




#22 AR-1248-2

R.T.: 6.281 min
Delta R.T.: 0.003 min
Response: 31620722
Conc: 507.37 ng/ml



#22 AR-1248-2


R.T.: 5.419 min
Delta R.T.: -0.006 min
Response: 33083239
Conc: 449.47 ng/ml



#23 AR-1248-3


R.T.: 6.489 min
Delta R.T.: 0.003 min
Response: 34347782
Conc: 510.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.463 min
Delta R.T.: -0.006 min
Response: 34441449
Conc: 450.70 ng/ml




#24 AR-1248-4

R.T.: 6.893 min
Delta R.T.: 0.003 min
Response: 37125233
Conc: 521.46 ng/ml



#24 AR-1248-4


R.T.: 5.639 min
Delta R.T.: -0.006 min
Response: 39596843
Conc: 450.40 ng/ml



#25 AR-1248-5

R.T.: 6.933 min
Delta R.T.: 0.002 min
Response: 36174651
Conc: 521.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 6.038 min
Delta R.T.: -0.007 min
Response: 35770657
Conc: 448.38 ng/ml