

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP062518.M
 Title : 8080.M
 Last Update : Tue Jun 26 06:52:58 2018
 Response Via : Initial Calibration

Calibration Files

750 =PP026648.D 1000 =PP026649.D 1500 =PP026650.D
 500 =PP026647.D 200 =PP026646.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	8.026	7.785	7.499	8.039	8.395	7.949	E5	4.18
2) T	3,5-DICHLOROBENZOI...	7.485	7.154	6.853	7.545	8.174	7.442	E5	6.65
3) T	4-Nitrophenol	3.611	3.455	3.293	3.667	4.005	3.606	E5	7.39
4) S	2,4-DCAA	4.955	4.784	4.543	5.029	5.482	4.959	E5	7.00
5) T	DICAMBA	2.100	2.025	1.937	2.116	2.160	2.068	E6	4.24
6) T	MCPP	1.329	1.311	1.272	1.286	1.207	1.281	E6	3.65
7) T	MCPA	2.036	2.041	1.944	2.019	2.012	2.010	E6	1.95
8) T	DICHLORPROP	5.259	5.020	4.782	5.461	6.075	5.320	E5	9.27
9) T	2,4-D	6.848	6.520	6.280	7.008	7.712	6.874	E5	7.96
10) T	Pentachlorophenol	8.446	7.852	7.174	8.764	9.291	8.306	E6	9.87
11) T	2,4,5-TP (SILVEX)	3.387	3.234	3.037	3.434	3.523	3.323	E6	5.75
12) T	2,4,5-T	3.193	3.070	2.891	3.238	3.324	3.143	E6	5.35
13) T	2,4-DB	3.925	3.750	3.617	3.900	4.121	3.863	E5	4.92
14) T	DINOSEB	2.391	2.274	2.122	2.384	2.443	2.323	E6	5.51
15) T	Picloram	4.692	4.469	4.225	4.675	4.695	4.551	E6	4.51
16) T	DCPA	4.063	3.863	3.524	4.110	4.190	3.950	E6	6.75

Signal #2 Calibration Files

750 =PP026648.D 1000 =PP026649.D 1500 =PP026650.D
 500 =PP026647.D 200 =PP026646.D

	Compound	750	1000	1500	500	200	Avg		%RSD
1) T	Dalapon	4.794	4.563	4.345	4.754	5.191	4.729	E5	6.63
2) T	3,5-DICHLOROBENZOI...	4.409	4.173	3.969	4.553	4.935	4.408	E5	8.39
3) T	4-Nitrophenol	2.236	2.147	2.027	2.318	2.507	2.247	E5	8.07
4) S	2,4-DCAA	3.423	3.243	3.076	3.568	3.812	3.424	E5	8.32
5) T	DICAMBA	1.296	1.240	1.183	1.315	1.352	1.277	E6	5.19
6) T	MCPP	8.452	8.134	7.817	8.492	8.086	8.196	E5	3.41
7) T	MCPA	1.441	1.358	1.276	1.516	1.728	1.464	E6	11.81
8) T	DICHLORPROP	3.309	3.147	2.947	3.424	3.762	3.318	E5	9.24
9) T	2,4-D	4.533	4.317	4.097	4.618	4.995	4.512	E5	7.47
10) T	Pentachlorophenol	5.295	4.943	4.505	5.526	5.737	5.201	E6	9.38
11) T	2,4,5-TP (SILVEX)	2.101	2.000	1.893	2.158	2.200	2.070	E6	6.01
12) T	2,4,5-T	1.814	1.735	1.642	1.849	1.838	1.776	E6	4.91
13) T	2,4-DB	2.493	2.417	2.301	2.614	2.742	2.514	E5	6.81
14) T	DINOSEB	1.430	1.359	1.290	1.454	1.520	1.411	E6	6.30
15) T	Picloram	2.846	2.739	2.596	2.893	2.862	2.787	E6	4.36
16) T	DCPA	2.624	2.467	2.290	2.698	2.782	2.572	E6	7.62

(#) = Out of Range