

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062521\
 Data File : PP036794.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 Jun 2021 10:39
 Operator : DD\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 28 04:13:56 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 18 06:06:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.971	3.956	2279192	1575352	51.567	50.862
2)	SA Decachlor...	11.003	9.257	2268645	1469989	52.885	45.965
Target Compounds							
26)	L6 AR-1254-1	7.199	6.080	852495	789307	484.741	465.842
27)	L6 AR-1254-2	7.428	6.240	1286868	682667	475.611	458.966
28)	L6 AR-1254-3	7.810	6.660	1377269	1117217	464.972	459.539
29)	L6 AR-1254-4	8.105	6.900	1053048	699527	479.018	452.044
30)	L6 AR-1254-5	8.533	7.330	1146756	973318	482.207	452.646



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062521\
Data File : PP036794.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 Jun 2021 10:39
Operator : DD\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 28 04:13:56 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061721.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 18 06:06:19 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.971 min
Delta R.T.: 0.006 min
Response: 2279192
Conc: 51.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.956 min
Delta R.T.: -0.003 min
Response: 1575352
Conc: 50.86 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.003 min
Delta R.T.: 0.006 min
Response: 2268645
Conc: 52.89 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.257 min
Delta R.T.: -0.006 min
Response: 1469989
Conc: 45.97 ng/ml



#26 AR-1254-1


R.T.: 7.199 min
Delta R.T.: 0.004 min
Response: 852495
Conc: 484.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 6.080 min
Delta R.T.: -0.004 min
Response: 789307
Conc: 465.84 ng/ml




#27 AR-1254-2

R.T.: 7.428 min
Delta R.T.: 0.004 min
Response: 1286868
Conc: 475.61 ng/ml



#27 AR-1254-2


R.T.: 6.240 min
Delta R.T.: -0.005 min
Response: 682667
Conc: 458.97 ng/ml



#28 AR-1254-3


R.T.: 7.810 min
Delta R.T.: 0.004 min
Response: 1377269
Conc: 464.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.660 min
Delta R.T.: -0.005 min
Response: 1117217
Conc: 459.54 ng/ml




#29 AR-1254-4

R.T.: 8.105 min
Delta R.T.: 0.004 min
Response: 1053048
Conc: 479.02 ng/ml



#29 AR-1254-4


R.T.: 6.900 min
Delta R.T.: -0.005 min
Response: 699527
Conc: 452.04 ng/ml



#30 AR-1254-5

R.T.: 8.533 min
Delta R.T.: 0.004 min
Response: 1146756
Conc: 482.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 7.330 min
Delta R.T.: -0.005 min
Response: 973318
Conc: 452.65 ng/ml