

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062524\
 Data File : PP066020.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jun 2024 16:47
 Operator : YP\AJ
 Sample : P3021-04RE
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ARS20-0018RE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 26 00:35:46 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 06 05:25:59 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.827	4.065	36772128	49959585	19.385	17.892
2)	SA Decachlor...	10.845	9.245	60830666	65235866	30.928	44.282 #
Target Compounds							
26)	L6 AR-1254-1	6.865	5.994	7057613	12569555	90.092	102.012
27)	L6 AR-1254-2	7.085	6.142	14658922	9817349	124.090	93.239
28)	L6 AR-1254-3	7.454	6.554	18697591	26307578	153.217	166.247
29)	L6 AR-1254-4	7.737	6.781	15700886	13652830	176.560	147.858
30)	L6 AR-1254-5	8.161	7.202	51488699	42568183	525.619	332.589 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062524\
Data File : PP066020.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jun 2024 16:47
Operator : YP\AJ
Sample : P3021-04RE
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ARS20-0018RE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 26 00:35:46 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 06 05:25:59 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.827 min
Delta R.T.: 0.000 min
Response: 36772128
Conc: 19.39 ng/ml

Instrument :
ECD_P
ClientSampleId :
ARS20-0018RE




#1 Tetrachloro-m-xylene

R.T.: 4.065 min
Delta R.T.: -0.006 min
Response: 49959585
Conc: 17.89 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.845 min
Delta R.T.: 0.002 min
Response: 60830666
Conc: 30.93 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.245 min
Delta R.T.: -0.016 min
Response: 65235866
Conc: 44.28 ng/ml



#26 AR-1254-1


R.T.: 6.865 min
Delta R.T.: 0.000 min
Response: 7057613
Conc: 90.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
ARS20-0018RE




#26 AR-1254-1

R.T.: 5.994 min
Delta R.T.: -0.009 min
Response: 12569555
Conc: 102.01 ng/ml




#27 AR-1254-2

R.T.: 7.085 min
Delta R.T.: 0.003 min
Response: 14658922
Conc: 124.09 ng/ml



#27 AR-1254-2


R.T.: 6.142 min
Delta R.T.: -0.007 min
Response: 9817349
Conc: 93.24 ng/ml



#28 AR-1254-3


R.T.: 7.454 min
Delta R.T.: 0.002 min
Response: 18697591
Conc: 153.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
ARS20-0018RE




#28 AR-1254-3

R.T.: 6.554 min
Delta R.T.: -0.008 min
Response: 26307578
Conc: 166.25 ng/ml




#29 AR-1254-4

R.T.: 7.737 min
Delta R.T.: 0.001 min
Response: 15700886
Conc: 176.56 ng/ml



#29 AR-1254-4


R.T.: 6.781 min
Delta R.T.: -0.008 min
Response: 13652830
Conc: 147.86 ng/ml



#30 AR-1254-5

R.T.: 8.161 min
Delta R.T.: 0.004 min
Response: 51488699
Conc: 525.62 ng/ml

Instrument :
ECD_P
ClientSampleId :
ARS20-0018RE



#30 AR-1254-5

R.T.: 7.202 min
Delta R.T.: -0.010 min
Response: 42568183
Conc: 332.59 ng/ml