

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062724\
 Data File : PP066071.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jun 2024 12:05
 Operator : YP\AJ
 Sample : P3072-10
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 446

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/28/2024
 Supervised By :Ankita Jodhani 06/28/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 28 00:16:04 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 06 05:25:59 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.826	4.064	31566597	42308533	16.641	15.152
2) SA Decachlor...	10.846	9.246	32993060	20941972	16.774	14.215
Target Compounds						
16) L4 AR-1242-1	6.004	5.177	9932273	11654080	181.633m	162.177
17) L4 AR-1242-2	6.028	5.197	13863258	13473154	174.817m	135.481
18) L4 AR-1242-3	6.091	5.381	6505133	6702340	128.692m	124.605
19) L4 AR-1242-4	6.188	5.459	7044621	3948129	171.672m	74.129 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062724\
Data File : PP066071.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jun 2024 12:05
Operator : YP\AJ
Sample : P3072-10
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
446


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/28/2024
Supervised By :Ankita Jodhani 06/28/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 28 00:16:04 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 06 05:25:59 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.826 min
Delta R.T.: 0.000 min
Response: 31566597
Conc: 16.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
446


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/28/2024
Supervised By :Ankita Jodhani 06/28/2024




#1 Tetrachloro-m-xylene

R.T.: 4.064 min
Delta R.T.: -0.007 min
Response: 42308533
Conc: 15.15 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.846 min
Delta R.T.: 0.003 min
Response: 32993060
Conc: 16.77 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.246 min
Delta R.T.: -0.014 min
Response: 20941972
Conc: 14.22 ng/ml




#16 AR-1242-1

R.T.: 6.004 min
Delta R.T.: 0.000 min
Response: 9932273
Conc: 181.63 ng/ml

Instrument :
ECD_P
ClientSampleId :
446


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/28/2024
Supervised By :Ankita Jodhani 06/28/2024




#16 AR-1242-1

R.T.: 5.177 min
Delta R.T.: -0.009 min
Response: 11654080
Conc: 162.18 ng/ml




#17 AR-1242-2

R.T.: 6.028 min
Delta R.T.: 0.000 min
Response: 13863258
Conc: 174.82 ng/ml m



#17 AR-1242-2

R.T.: 5.197 min
Delta R.T.: -0.010 min
Response: 13473154
Conc: 135.48 ng/ml




#18 AR-1242-3

R.T.: 6.091 min
Delta R.T.: 0.000 min
Response: 6505133
Conc: 128.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
446


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/28/2024
Supervised By :Ankita Jodhani 06/28/2024




#18 AR-1242-3

R.T.: 5.381 min
Delta R.T.: -0.007 min
Response: 6702340
Conc: 124.60 ng/ml



#19 AR-1242-4

R.T.: 6.188 min
Delta R.T.: -0.003 min
Response: 7044621
Conc: 171.67 ng/ml m



#19 AR-1242-4

R.T.: 5.459 min
Delta R.T.: -0.011 min
Response: 3948129
Conc: 74.13 ng/ml