

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062724\
 Data File : PP066085.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jun 2024 16:26
 Operator : YP\AJ
 Sample : P3070-01DL 2X
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 624DL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 28 00:20:18 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 06 05:25:59 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.828	4.064	23684993	29957227	12.486m	10.728
2) SA Decachlor...	10.849	9.247	18662097	13089440	9.488	8.885
Target Compounds						
16) L4 AR-1242-1	6.005	5.177	58161458	51176707	1063.610m	712.168m#
17) L4 AR-1242-2	6.030	5.197	58675643	74281299	739.907m	746.946m
18) L4 AR-1242-3	6.092	5.381	42995422	29942911	850.582m	556.675m#
19) L4 AR-1242-4	6.192	5.461	34250653	25013374	834.660m	469.645m#
20) L4 AR-1242-5	6.933	5.994	11367047	16087431	260.581	270.749



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062724\
Data File : PP066085.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jun 2024 16:26
Operator : YP\AJ
Sample : P3070-01DL 2X
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
624DL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 28 00:20:18 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP060524.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 06 05:25:59 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.828 min
Delta R.T.: 0.002 min
Response: 23684993
Conc: 12.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
624DL




#1 Tetrachloro-m-xylene

R.T.: 4.064 min
Delta R.T.: -0.007 min
Response: 29957227
Conc: 10.73 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.849 min
Delta R.T.: 0.006 min
Response: 18662097
Conc: 9.49 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.247 min
Delta R.T.: -0.013 min
Response: 13089440
Conc: 8.89 ng/ml



#16 AR-1242-1


R.T.: 6.005 min
Delta R.T.: 0.000 min
Response: 58161458
Conc: 1063.61 ng/ml

Instrument :
ECD_P
ClientSampleId :
624DL




#16 AR-1242-1

R.T.: 5.177 min
Delta R.T.: -0.009 min
Response: 51176707
Conc: 712.17 ng/ml m




#17 AR-1242-2

R.T.: 6.030 min
Delta R.T.: 0.002 min
Response: 58675643
Conc: 739.91 ng/ml m



#17 AR-1242-2


R.T.: 5.197 min
Delta R.T.: -0.009 min
Response: 74281299
Conc: 746.95 ng/ml m



#18 AR-1242-3


R.T.: 6.092 min
Delta R.T.: 0.001 min
Response: 42995422
Conc: 850.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
624DL




#18 AR-1242-3

R.T.: 5.381 min
Delta R.T.: -0.008 min
Response: 29942911
Conc: 556.67 ng/ml m




#19 AR-1242-4

R.T.: 6.192 min
Delta R.T.: 0.001 min
Response: 34250653
Conc: 834.66 ng/ml m



#19 AR-1242-4


R.T.: 5.461 min
Delta R.T.: -0.009 min
Response: 25013374
Conc: 469.64 ng/ml m



#20 AR-1242-5

R.T.: 6.933 min
Delta R.T.: 0.003 min
Response: 11367047
Conc: 260.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
624DL



#20 AR-1242-5

R.T.: 5.994 min
Delta R.T.: -0.008 min
Response: 16087431
Conc: 270.75 ng/ml