

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062725\
 Data File : PP073336.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jun 2025 18:57
 Operator : YP\AJ
 Sample : Q2436-02
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 TP-69

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/30/2025
 Supervised By :mohammad ahmed 07/01/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 27 23:22:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061725.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 18 08:17:15 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.488	3.783	33572685	30174904	15.905	16.747
2) SA Decachlor...	10.175	8.787	20496839	18995155	11.994m	14.232m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062725\
Data File : PP073336.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jun 2025 18:57
Operator : YP\AJ
Sample : Q2436-02
Misc :
ALS Vial : 17 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
TP-69


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/30/2025
Supervised By :mohammad ahmed 07/01/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 27 23:22:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061725.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 18 08:17:15 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.488 min
Delta R.T.: -0.006 min
Response: 33572685
Conc: 15.91 ng/ml

Instrument :
ECD_P
ClientSampleId :
TP-69


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/30/2025
Supervised By :mohammad ahmed 07/01/2025




#1 Tetrachloro-m-xylene

R.T.: 3.783 min
Delta R.T.: -0.005 min
Response: 30174904
Conc: 16.75 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.175 min
Delta R.T.: -0.013 min
Response: 20496839
Conc: 11.99 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.787 min
Delta R.T.: -0.010 min
Response: 18995155
Conc: 14.23 ng/ml m