

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062821\
 Data File : PP036817.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Jun 2021 9:48
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 29 06:53:32 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 18 06:06:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.965	3.955	2048448	1412829	46.346	45.614
2) SA Decachlor...	10.995	9.250	1961941	1280055	45.736	40.026
Target Compounds						
3) L1 AR-1016-1	6.284	5.206	757734	510237	447.344	432.823
4) L1 AR-1016-2	6.307	5.226	1088758	710075	451.533	437.201
5) L1 AR-1016-3	6.374	5.416	691398	391360	449.762	433.776
6) L1 AR-1016-4	6.481	5.470	579455	300941	450.746	423.931
7) L1 AR-1016-5	6.795	5.697	569716	390868	445.420	431.314
31) L7 AR-1260-1	7.970	6.794	916610	662079	422.303	411.219
32) L7 AR-1260-2	8.235	6.993	1114872	805580	434.996	415.405
33) L7 AR-1260-3	8.601	7.147	865896	746714	431.980	410.577
34) L7 AR-1260-4	8.831	7.630	1038537	640157	438.342	412.806
35) L7 AR-1260-5	9.160	7.879	1955443	1504828	438.252	409.505



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062821\
Data File : PP036817.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 Jun 2021 9:48
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 29 06:53:32 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061721.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 18 06:06:19 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.965 min
Delta R.T.: 0.000 min
Response: 2048448
Conc: 46.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.955 min
Delta R.T.: -0.004 min
Response: 1412829
Conc: 45.61 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.995 min
Delta R.T.: -0.003 min
Response: 1961941
Conc: 45.74 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.250 min
Delta R.T.: -0.013 min
Response: 1280055
Conc: 40.03 ng/ml



#3 AR-1016-1


R.T.: 6.284 min
Delta R.T.: 0.000 min
Response: 757734
Conc: 447.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 5.206 min
Delta R.T.: -0.007 min
Response: 510237
Conc: 432.82 ng/ml




#4 AR-1016-2

R.T.: 6.307 min
Delta R.T.: -0.002 min
Response: 1088758
Conc: 451.53 ng/ml



#4 AR-1016-2


R.T.: 5.226 min
Delta R.T.: -0.007 min
Response: 710075
Conc: 437.20 ng/ml



#5 AR-1016-3


R.T.: 6.374 min
Delta R.T.: 0.000 min
Response: 691398
Conc: 449.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.416 min
Delta R.T.: -0.008 min
Response: 391360
Conc: 433.78 ng/ml




#6 AR-1016-4

R.T.: 6.481 min
Delta R.T.: 0.000 min
Response: 579455
Conc: 450.75 ng/ml



#6 AR-1016-4


R.T.: 5.470 min
Delta R.T.: -0.007 min
Response: 300941
Conc: 423.93 ng/ml



#7 AR-1016-5


R.T.: 6.795 min
Delta R.T.: -0.002 min
Response: 569716
Conc: 445.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.697 min
Delta R.T.: -0.008 min
Response: 390868
Conc: 431.31 ng/ml




#31 AR-1260-1

R.T.: 7.970 min
Delta R.T.: -0.001 min
Response: 916610
Conc: 422.30 ng/ml



#31 AR-1260-1


R.T.: 6.794 min
Delta R.T.: -0.009 min
Response: 662079
Conc: 411.22 ng/ml



#32 AR-1260-2


R.T.: 8.235 min
Delta R.T.: -0.002 min
Response: 1114872
Conc: 435.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.993 min
Delta R.T.: -0.009 min
Response: 805580
Conc: 415.41 ng/ml




#33 AR-1260-3

R.T.: 8.601 min
Delta R.T.: -0.002 min
Response: 865896
Conc: 431.98 ng/ml



#33 AR-1260-3


R.T.: 7.147 min
Delta R.T.: -0.008 min
Response: 746714
Conc: 410.58 ng/ml



#34 AR-1260-4


R.T.: 8.831 min
Delta R.T.: -0.003 min
Response: 1038537
Conc: 438.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.630 min
Delta R.T.: -0.010 min
Response: 640157
Conc: 412.81 ng/ml



#35 AR-1260-5

R.T.: 9.160 min
Delta R.T.: -0.003 min
Response: 1955443
Conc: 438.25 ng/ml



#35 AR-1260-5

R.T.: 7.879 min
Delta R.T.: -0.009 min
Response: 1504828
Conc: 409.50 ng/ml