

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062822\
 Data File : PP048984.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jun 2022 02:29
 Operator : YP\AJ
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 29 03:50:30 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 29 03:43:47 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.345	3.594	106.7E6	188.5E6	74.346	74.291
2) SA Decachlor...	10.095	8.567	91759005	79979562	74.727	74.257
Target Compounds						
3) L1 AR-1016-1	5.523	4.670	37733551	64608119	739.760	740.527
4) L1 AR-1016-2	5.545	4.689	54173599	92663280	744.161	739.208
5) L1 AR-1016-3	5.607	4.864	33518719	49214108	746.580	741.977
6) L1 AR-1016-4	5.707	4.905	28005212	38239372	744.490	740.000
7) L1 AR-1016-5	6.002	5.117	27549856	49771078	744.532	731.553
31) L7 AR-1260-1	7.130	6.145	52521636	79110842	747.300	738.438
32) L7 AR-1260-2	7.388	6.332	61537806	91284301	747.428	740.714
33) L7 AR-1260-3	7.746	6.485	41519635	84225446	747.269	739.578
34) L7 AR-1260-4	7.974	6.956	51641056	58763471	745.956	738.825
35) L7 AR-1260-5	8.294	7.196	102.1E6	133.3E6	747.131	743.970



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062822\
Data File : PP048984.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jun 2022 02:29
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 29 03:50:30 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 29 03:43:47 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.345 min
Delta R.T.: 0.000 min
Response: 106711518
Conc: 74.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.594 min
Delta R.T.: 0.000 min
Response: 188500791
Conc: 74.29 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.095 min
Delta R.T.: 0.000 min
Response: 91759005
Conc: 74.73 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.567 min
Delta R.T.: 0.000 min
Response: 79979562
Conc: 74.26 ng/ml



#3 AR-1016-1


R.T.: 5.523 min
Delta R.T.: 0.000 min
Response: 37733551
Conc: 739.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#3 AR-1016-1

R.T.: 4.670 min
Delta R.T.: 0.000 min
Response: 64608119
Conc: 740.53 ng/ml




#4 AR-1016-2

R.T.: 5.545 min
Delta R.T.: 0.000 min
Response: 54173599
Conc: 744.16 ng/ml



#4 AR-1016-2


R.T.: 4.689 min
Delta R.T.: 0.000 min
Response: 92663280
Conc: 739.21 ng/ml



#5 AR-1016-3


R.T.: 5.607 min
Delta R.T.: 0.000 min
Response: 33518719
Conc: 746.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#5 AR-1016-3

R.T.: 4.864 min
Delta R.T.: 0.000 min
Response: 49214108
Conc: 741.98 ng/ml




#6 AR-1016-4

R.T.: 5.707 min
Delta R.T.: 0.000 min
Response: 28005212
Conc: 744.49 ng/ml



#6 AR-1016-4


R.T.: 4.905 min
Delta R.T.: 0.000 min
Response: 38239372
Conc: 740.00 ng/ml



#7 AR-1016-5


R.T.: 6.002 min
Delta R.T.: 0.000 min
Response: 27549856
Conc: 744.53 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#7 AR-1016-5

R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 49771078
Conc: 731.55 ng/ml




#31 AR-1260-1

R.T.: 7.130 min
Delta R.T.: 0.000 min
Response: 52521636
Conc: 747.30 ng/ml



#31 AR-1260-1


R.T.: 6.145 min
Delta R.T.: 0.000 min
Response: 79110842
Conc: 738.44 ng/ml



#32 AR-1260-2


R.T.: 7.388 min
Delta R.T.: 0.000 min
Response: 61537806
Conc: 747.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#32 AR-1260-2

R.T.: 6.332 min
Delta R.T.: 0.000 min
Response: 91284301
Conc: 740.71 ng/ml




#33 AR-1260-3

R.T.: 7.746 min
Delta R.T.: 0.000 min
Response: 41519635
Conc: 747.27 ng/ml



#33 AR-1260-3


R.T.: 6.485 min
Delta R.T.: 0.000 min
Response: 84225446
Conc: 739.58 ng/ml



#34 AR-1260-4


R.T.: 7.974 min
Delta R.T.: 0.000 min
Response: 51641056
Conc: 745.96 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#34 AR-1260-4

R.T.: 6.956 min
Delta R.T.: 0.000 min
Response: 58763471
Conc: 738.82 ng/ml



#35 AR-1260-5

R.T.: 8.294 min
Delta R.T.: 0.000 min
Response: 102117691
Conc: 747.13 ng/ml



#35 AR-1260-5

R.T.: 7.196 min
Delta R.T.: 0.000 min
Response: 133302238
Conc: 743.97 ng/ml