

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062922\
 Data File : PP049012.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jun 2022 11:34
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 29 18:41:42 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 29 09:37:18 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.345	3.593	72257565	127.6E6	50.829	49.834
2)	SA Decachlor...	10.095	8.567	62735168	56455973	51.146	50.085
Target Compounds							
3)	L1 AR-1016-1	5.523	4.669	26740548	44819218	508.642	491.065
4)	L1 AR-1016-2	5.546	4.688	37451670	63538483	496.970	493.728
5)	L1 AR-1016-3	5.607	4.864	23287105	34093759	478.354	460.742
6)	L1 AR-1016-4	5.708	4.904	19463302	27132481	501.422	530.362
7)	L1 AR-1016-5	6.002	5.117	19641414	35207851	514.804	471.965
8)	L2 AR-1221-1	4.554	3.804	2327192	4787393	136.514	142.631
9)	L2 AR-1221-2	4.646	3.890	3534306	6968465	276.461	271.454
10)	L2 AR-1221-3	4.723	3.965	13766398	24976143	347.864	339.651
11)	L3 AR-1232-1	4.723	3.965	13766398	24976143	362.064	359.513
12)	L3 AR-1232-2	5.253	4.688	18893808	63538483	999.557	988.960
13)	L3 AR-1232-3	5.546	4.864	37451670	34093759	1013.277	968.416
14)	L3 AR-1232-4	5.708	4.946	19463302	33633282	1018.236	1062.785
15)	L3 AR-1232-5	5.796	5.117	17065713	35207851	1094.303	983.427
16)	L4 AR-1242-1	5.523	4.669	26740548	44819218	659.879	623.185
17)	L4 AR-1242-2	5.546	4.688	37451670	63538483	650.292	627.347
18)	L4 AR-1242-3	5.607	4.864	23287105	34093759	643.205	611.603
19)	L4 AR-1242-4	5.708	4.946	19463302	33633282	661.587	607.547
20)	L4 AR-1242-5	6.447	5.467	2685341	26485466	80.471	380.424 #
21)	L5 AR-1248-1	5.523	4.669	26740548	44819218	831.522	771.116
22)	L5 AR-1248-2	5.796	4.904	17065713	27132481	367.106	404.297
23)	L5 AR-1248-3	6.002	4.946	19641414	33633282	366.791	410.370
24)	L5 AR-1248-4	6.408	5.117	3287564	35207851	53.196	356.898 #
25)	L5 AR-1248-5	6.447	5.507	2685341	4752483	46.577	47.377
26)	L6 AR-1254-1	6.380	5.467	14538031	26485466	215.670	180.407
27)	L6 AR-1254-2	6.598	5.613	15821698	24521266	168.695	196.550
28)	L6 AR-1254-3	6.969	6.030	8973239	41682939	88.433	231.500 #
29)	L6 AR-1254-4	7.255	6.243	5924423	5313238	78.966	49.111 #
30)	L6 AR-1254-5	7.675	6.659	52286748	67918309	619.746	467.950
31)	L7 AR-1260-1	7.130	6.145	35992913	54677489	494.762	474.616
32)	L7 AR-1260-2	7.389	6.332	42807952	62748167	508.624	467.813
33)	L7 AR-1260-3	7.747	6.486	29084292	58140987	519.865	477.611
34)	L7 AR-1260-4	7.975	6.956	35966080	41036147	514.290	486.373
35)	L7 AR-1260-5	8.295	7.196	69222210	90451058	505.190	479.788
36)	L8 AR-1262-1	7.747	6.696	29084292	45912718	310.294	345.146
37)	L8 AR-1262-2	8.295	6.956	69222210	41036147	404.512	362.127
38)	L8 AR-1262-3	8.625f	7.479	46065245	19264161	615.580	236.434 #
39)	L8 AR-1262-4	8.674f	7.542	26678659	62256368	568.399	427.094
40)	L8 AR-1262-5	9.353	8.035	18098719	18730538	299.346	310.539
41)	L9 AR-1268-1	8.625f	7.479	46065245	19264161	245.291	89.549 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062922\
 Data File : PP049012.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jun 2022 11:34
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 29 18:41:42 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 29 09:37:18 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.674f	7.542	26678659	62256368	152.483	325.458 #
43) L9	AR-1268-3	8.918	7.749	1401459	1491837	9.783	9.795
44) L9	AR-1268-4	9.353	8.035	18098719	18730538	282.527	293.898
45) L9	AR-1268-5	9.763	8.318	5457740	5506220	11.833	13.800



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062922\
Data File : PP049012.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jun 2022 11:34
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 29 18:41:22 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 29 09:37:18 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.345 min
Delta R.T.: 0.000 min
Response: 72257565
Conc: 50.83 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.593 min
Delta R.T.: 0.000 min
Response: 127607745
Conc: 49.83 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.095 min
Delta R.T.: 0.001 min
Response: 62735168
Conc: 51.15 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.567 min
Delta R.T.: 0.000 min
Response: 56455973
Conc: 50.08 ng/ml



#3 AR-1016-1


R.T.: 5.523 min
Delta R.T.: 0.000 min
Response: 26740548
Conc: 508.64 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 4.669 min
Delta R.T.: 0.000 min
Response: 44819218
Conc: 491.07 ng/ml




#4 AR-1016-2

R.T.: 5.546 min
Delta R.T.: 0.000 min
Response: 37451670
Conc: 496.97 ng/ml



#4 AR-1016-2


R.T.: 4.688 min
Delta R.T.: 0.000 min
Response: 63538483
Conc: 493.73 ng/ml



#5 AR-1016-3


R.T.: 5.607 min
Delta R.T.: 0.000 min
Response: 23287105
Conc: 478.35 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 4.864 min
Delta R.T.: 0.000 min
Response: 34093759
Conc: 460.74 ng/ml




#6 AR-1016-4

R.T.: 5.708 min
Delta R.T.: 0.000 min
Response: 19463302
Conc: 501.42 ng/ml



#6 AR-1016-4


R.T.: 4.904 min
Delta R.T.: 0.000 min
Response: 27132481
Conc: 530.36 ng/ml



#7 AR-1016-5


R.T.: 6.002 min
Delta R.T.: 0.000 min
Response: 19641414
Conc: 514.80 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 35207851
Conc: 471.96 ng/ml




#8 AR-1221-1

R.T.: 4.554 min
Delta R.T.: -0.005 min
Response: 2327192
Conc: 136.51 ng/ml



#8 AR-1221-1


R.T.: 3.804 min
Delta R.T.: -0.002 min
Response: 4787393
Conc: 142.63 ng/ml



#9 AR-1221-2


R.T.: 4.646 min
Delta R.T.: 0.000 min
Response: 3534306
Conc: 276.46 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 3.890 min
Delta R.T.: 0.000 min
Response: 6968465
Conc: 271.45 ng/ml




#10 AR-1221-3

R.T.: 4.723 min
Delta R.T.: 0.000 min
Response: 13766398
Conc: 347.86 ng/ml



#10 AR-1221-3


R.T.: 3.965 min
Delta R.T.: 0.000 min
Response: 24976143
Conc: 339.65 ng/ml



#11 AR-1232-1


R.T.: 4.723 min
Delta R.T.: 0.000 min
Response: 13766398
Conc: 362.06 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 3.965 min
Delta R.T.: -0.001 min
Response: 24976143
Conc: 359.51 ng/ml




#12 AR-1232-2

R.T.: 5.253 min
Delta R.T.: 0.000 min
Response: 18893808
Conc: 999.56 ng/ml



#12 AR-1232-2


R.T.: 4.688 min
Delta R.T.: 0.000 min
Response: 63538483
Conc: 988.96 ng/ml



#13 AR-1232-3


R.T.: 5.546 min
Delta R.T.: 0.000 min
Response: 37451670
Conc: 1013.28 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 4.864 min
Delta R.T.: 0.000 min
Response: 34093759
Conc: 968.42 ng/ml




#14 AR-1232-4

R.T.: 5.708 min
Delta R.T.: -0.001 min
Response: 19463302
Conc: 1018.24 ng/ml



#14 AR-1232-4


R.T.: 4.946 min
Delta R.T.: 0.000 min
Response: 33633282
Conc: 1062.79 ng/ml



#15 AR-1232-5


R.T.: 5.796 min
Delta R.T.: 0.000 min
Response: 17065713
Conc: 1094.30 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 35207851
Conc: 983.43 ng/ml




#16 AR-1242-1

R.T.: 5.523 min
Delta R.T.: 0.000 min
Response: 26740548
Conc: 659.88 ng/ml



#16 AR-1242-1


R.T.: 4.669 min
Delta R.T.: 0.000 min
Response: 44819218
Conc: 623.18 ng/ml



#17 AR-1242-2


R.T.: 5.546 min
Delta R.T.: 0.000 min
Response: 37451670
Conc: 650.29 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 4.688 min
Delta R.T.: 0.000 min
Response: 63538483
Conc: 627.35 ng/ml




#18 AR-1242-3

R.T.: 5.607 min
Delta R.T.: 0.000 min
Response: 23287105
Conc: 643.20 ng/ml



#18 AR-1242-3


R.T.: 4.864 min
Delta R.T.: 0.000 min
Response: 34093759
Conc: 611.60 ng/ml



#19 AR-1242-4


R.T.: 5.708 min
Delta R.T.: 0.000 min
Response: 19463302
Conc: 661.59 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 4.946 min
Delta R.T.: 0.000 min
Response: 33633282
Conc: 607.55 ng/ml




#20 AR-1242-5

R.T.: 6.447 min
Delta R.T.: 0.000 min
Response: 2685341
Conc: 80.47 ng/ml



#20 AR-1242-5


R.T.: 5.467 min
Delta R.T.: 0.000 min
Response: 26485466
Conc: 380.42 ng/ml



#21 AR-1248-1


R.T.: 5.523 min
Delta R.T.: 0.001 min
Response: 26740548
Conc: 831.52 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 4.669 min
Delta R.T.: 0.000 min
Response: 44819218
Conc: 771.12 ng/ml




#22 AR-1248-2

R.T.: 5.796 min
Delta R.T.: 0.000 min
Response: 17065713
Conc: 367.11 ng/ml



#22 AR-1248-2


R.T.: 4.904 min
Delta R.T.: 0.000 min
Response: 27132481
Conc: 404.30 ng/ml



#23 AR-1248-3


R.T.: 6.002 min
Delta R.T.: 0.001 min
Response: 19641414
Conc: 366.79 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 4.946 min
Delta R.T.: 0.000 min
Response: 33633282
Conc: 410.37 ng/ml




#24 AR-1248-4

R.T.: 6.408 min
Delta R.T.: 0.000 min
Response: 3287564
Conc: 53.20 ng/ml



#24 AR-1248-4


R.T.: 5.117 min
Delta R.T.: 0.000 min
Response: 35207851
Conc: 356.90 ng/ml



#25 AR-1248-5


R.T.: 6.447 min
Delta R.T.: 0.000 min
Response: 2685341
Conc: 46.58 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 5.507 min
Delta R.T.: 0.000 min
Response: 4752483
Conc: 47.38 ng/ml




#26 AR-1254-1

R.T.: 6.380 min
Delta R.T.: 0.001 min
Response: 14538031
Conc: 215.67 ng/ml



#26 AR-1254-1


R.T.: 5.467 min
Delta R.T.: 0.000 min
Response: 26485466
Conc: 180.41 ng/ml



#27 AR-1254-2


R.T.: 6.598 min
Delta R.T.: 0.000 min
Response: 15821698
Conc: 168.70 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 5.613 min
Delta R.T.: 0.000 min
Response: 24521266
Conc: 196.55 ng/ml




#28 AR-1254-3

R.T.: 6.969 min
Delta R.T.: 0.000 min
Response: 8973239
Conc: 88.43 ng/ml



#28 AR-1254-3


R.T.: 6.030 min
Delta R.T.: 0.014 min
Response: 41682939
Conc: 231.50 ng/ml



#29 AR-1254-4


R.T.: 7.255 min
Delta R.T.: 0.000 min
Response: 5924423
Conc: 78.97 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.243 min
Delta R.T.: 0.000 min
Response: 5313238
Conc: 49.11 ng/ml




#30 AR-1254-5

R.T.: 7.675 min
Delta R.T.: 0.000 min
Response: 52286748
Conc: 619.75 ng/ml



#30 AR-1254-5


R.T.: 6.659 min
Delta R.T.: -0.001 min
Response: 67918309
Conc: 467.95 ng/ml



#31 AR-1260-1


R.T.: 7.130 min
Delta R.T.: 0.000 min
Response: 35992913
Conc: 494.76 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.145 min
Delta R.T.: 0.000 min
Response: 54677489
Conc: 474.62 ng/ml




#32 AR-1260-2

R.T.: 7.389 min
Delta R.T.: 0.000 min
Response: 42807952
Conc: 508.62 ng/ml



#32 AR-1260-2


R.T.: 6.332 min
Delta R.T.: 0.000 min
Response: 62748167
Conc: 467.81 ng/ml



#33 AR-1260-3


R.T.: 7.747 min
Delta R.T.: 0.000 min
Response: 29084292
Conc: 519.86 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 6.486 min
Delta R.T.: 0.000 min
Response: 58140987
Conc: 477.61 ng/ml




#34 AR-1260-4

R.T.: 7.975 min
Delta R.T.: 0.000 min
Response: 35966080
Conc: 514.29 ng/ml



#34 AR-1260-4


R.T.: 6.956 min
Delta R.T.: 0.000 min
Response: 41036147
Conc: 486.37 ng/ml



#35 AR-1260-5


R.T.: 8.295 min
Delta R.T.: 0.000 min
Response: 69222210
Conc: 505.19 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.196 min
Delta R.T.: 0.000 min
Response: 90451058
Conc: 479.79 ng/ml




#36 AR-1262-1

R.T.: 7.747 min
Delta R.T.: 0.002 min
Response: 29084292
Conc: 310.29 ng/ml



#36 AR-1262-1


R.T.: 6.696 min
Delta R.T.: 0.000 min
Response: 45912718
Conc: 345.15 ng/ml



#37 AR-1262-2


R.T.: 8.295 min
Delta R.T.: 0.002 min
Response: 69222210
Conc: 404.51 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 6.956 min
Delta R.T.: 0.000 min
Response: 41036147
Conc: 362.13 ng/ml




#38 AR-1262-3

R.T.: 8.625 min
Delta R.T.: 0.025 min
Response: 46065245
Conc: 615.58 ng/ml



#38 AR-1262-3


R.T.: 7.479 min
Delta R.T.: 0.001 min
Response: 19264161
Conc: 236.43 ng/ml



#39 AR-1262-4


R.T.: 8.674 min
Delta R.T.: -0.017 min
Response: 26678659
Conc: 568.40 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 7.542 min
Delta R.T.: 0.000 min
Response: 62256368
Conc: 427.09 ng/ml




#40 AR-1262-5

R.T.: 9.353 min
Delta R.T.: 0.002 min
Response: 18098719
Conc: 299.35 ng/ml



#40 AR-1262-5


R.T.: 8.035 min
Delta R.T.: 0.001 min
Response: 18730538
Conc: 310.54 ng/ml



#41 AR-1268-1


R.T.: 8.625 min
Delta R.T.: 0.026 min
Response: 46065245
Conc: 245.29 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 7.479 min
Delta R.T.: 0.001 min
Response: 19264161
Conc: 89.55 ng/ml




#42 AR-1268-2

R.T.: 8.674 min
Delta R.T.: -0.019 min
Response: 26678659
Conc: 152.48 ng/ml



#42 AR-1268-2


R.T.: 7.542 min
Delta R.T.: 0.000 min
Response: 62256368
Conc: 325.46 ng/ml



#43 AR-1268-3


R.T.: 8.918 min
Delta R.T.: 0.002 min
Response: 1401459
Conc: 9.78 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 7.749 min
Delta R.T.: 0.000 min
Response: 1491837
Conc: 9.79 ng/ml




#44 AR-1268-4

R.T.: 9.353 min
Delta R.T.: 0.003 min
Response: 18098719
Conc: 282.53 ng/ml



#44 AR-1268-4


R.T.: 8.035 min
Delta R.T.: 0.001 min
Response: 18730538
Conc: 293.90 ng/ml



#45 AR-1268-5

R.T.: 9.763 min
Delta R.T.: 0.004 min
Response: 5457740
Conc: 11.83 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 8.318 min
Delta R.T.: 0.000 min
Response: 5506220
Conc: 13.80 ng/ml