

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062922\
 Data File : PP049030.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jun 2022 17:07
 Operator : YP\AJ
 Sample : N3214-07
 Misc : AR1248 LOD 25 PPB
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 LOD-MDL-WATER-01-QT2-2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 29 18:50:44 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 29 09:37:18 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.344	3.593	26788256	49234298	18.844	19.227
2) SA Decachlor...	10.092	8.566	25487100	23347577	20.779	20.713
Target Compounds						
21) L5 AR-1248-1	5.521	4.669	696922	1326687	21.671	22.826
22) L5 AR-1248-2	5.794	4.904	1165091	1891637	25.063	28.187
23) L5 AR-1248-3	6.000	4.944	1232327	2046370	23.013	24.968
24) L5 AR-1248-4	6.407	5.115	1310599	2547666	21.207	25.825
25) L5 AR-1248-5	6.445	5.506	1191501	2749169	20.667	27.406 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062922\
Data File : PP049030.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jun 2022 17:07
Operator : YP\AJ
Sample : N3214-07
Misc : AR1248 LOD 25 PPB
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT2-2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 29 18:50:44 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 29 09:37:18 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.344 min
Delta R.T.: -0.002 min
Response: 26788256
Conc: 18.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT2-2022




#1 Tetrachloro-m-xylene

R.T.: 3.593 min
Delta R.T.: -0.001 min
Response: 49234298
Conc: 19.23 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.092 min
Delta R.T.: -0.003 min
Response: 25487100
Conc: 20.78 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.566 min
Delta R.T.: -0.002 min
Response: 23347577
Conc: 20.71 ng/ml



#21 AR-1248-1


R.T.: 5.521 min
Delta R.T.: 0.000 min
Response: 696922
Conc: 21.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT2-2022




#21 AR-1248-1

R.T.: 4.669 min
Delta R.T.: -0.001 min
Response: 1326687
Conc: 22.83 ng/ml




#22 AR-1248-2

R.T.: 5.794 min
Delta R.T.: 0.000 min
Response: 1165091
Conc: 25.06 ng/ml



#22 AR-1248-2


R.T.: 4.904 min
Delta R.T.: -0.001 min
Response: 1891637
Conc: 28.19 ng/ml



#23 AR-1248-3


R.T.: 6.000 min
Delta R.T.: 0.000 min
Response: 1232327
Conc: 23.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT2-2022




#23 AR-1248-3

R.T.: 4.944 min
Delta R.T.: -0.002 min
Response: 2046370
Conc: 24.97 ng/ml




#24 AR-1248-4

R.T.: 6.407 min
Delta R.T.: -0.002 min
Response: 1310599
Conc: 21.21 ng/ml



#24 AR-1248-4


R.T.: 5.115 min
Delta R.T.: -0.001 min
Response: 2547666
Conc: 25.83 ng/ml



#25 AR-1248-5

R.T.: 6.445 min
Delta R.T.: -0.001 min
Response: 1191501
Conc: 20.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT2-2022



#25 AR-1248-5

R.T.: 5.506 min
Delta R.T.: -0.002 min
Response: 2749169
Conc: 27.41 ng/ml