

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062922\
 Data File : PP049050.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jun 2022 22:52
 Operator : YP\AJ
 Sample : N3214-07
 Misc : AR1248 LOD 40 PPB
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 LOD-MDL-WATER-01-QT2-2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 30 08:08:19 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 29 09:37:18 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.344	3.593	28366515	51743776	19.954	20.207
2) SA Decachlor...	10.092	8.566	28266100	24372105	23.044	21.622
Target Compounds						
21) L5 AR-1248-1	5.521	4.669	1196967	2190336	37.221	37.685
22) L5 AR-1248-2	5.794	4.903	1983556	3179646	42.669	47.379
23) L5 AR-1248-3	5.999	4.945	2096351	3392838	39.148	41.397
24) L5 AR-1248-4	6.407	5.115	2314152	3960735	37.445	40.149
25) L5 AR-1248-5	6.446	5.507	2193732	3809149	38.050	37.973



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP062922\
Data File : PP049050.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jun 2022 22:52
Operator : YP\AJ
Sample : N3214-07
Misc : AR1248 LOD 40 PPB
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT2-2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 30 08:08:19 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 29 09:37:18 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.344 min
Delta R.T.: -0.002 min
Response: 28366515
Conc: 19.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT2-2022




#1 Tetrachloro-m-xylene

R.T.: 3.593 min
Delta R.T.: 0.000 min
Response: 51743776
Conc: 20.21 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.092 min
Delta R.T.: -0.002 min
Response: 28266100
Conc: 23.04 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.566 min
Delta R.T.: -0.002 min
Response: 24372105
Conc: 21.62 ng/ml



#21 AR-1248-1


R.T.: 5.521 min
Delta R.T.: 0.000 min
Response: 1196967
Conc: 37.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT2-2022




#21 AR-1248-1

R.T.: 4.669 min
Delta R.T.: -0.001 min
Response: 2190336
Conc: 37.68 ng/ml




#22 AR-1248-2

R.T.: 5.794 min
Delta R.T.: -0.001 min
Response: 1983556
Conc: 42.67 ng/ml



#22 AR-1248-2


R.T.: 4.903 min
Delta R.T.: -0.001 min
Response: 3179646
Conc: 47.38 ng/ml



#23 AR-1248-3


R.T.: 5.999 min
Delta R.T.: -0.001 min
Response: 2096351
Conc: 39.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT2-2022




#23 AR-1248-3

R.T.: 4.945 min
Delta R.T.: -0.001 min
Response: 3392838
Conc: 41.40 ng/ml




#24 AR-1248-4

R.T.: 6.407 min
Delta R.T.: -0.001 min
Response: 2314152
Conc: 37.45 ng/ml



#24 AR-1248-4


R.T.: 5.115 min
Delta R.T.: -0.002 min
Response: 3960735
Conc: 40.15 ng/ml



#25 AR-1248-5

R.T.: 6.446 min
Delta R.T.: 0.000 min
Response: 2193732
Conc: 38.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
LOD-MDL-WATER-01-QT2-2022



#25 AR-1248-5

R.T.: 5.507 min
Delta R.T.: 0.000 min
Response: 3809149
Conc: 37.97 ng/ml