

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP063021\
 Data File : PP036905.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Jun 2021 10:26
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 30 13:56:15 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 18 06:06:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.962	3.954	2579691	1704280	58.366	55.024
2)	SA Decachlor...	10.994	9.249	2408167	1463254	56.138	45.755
Target Compounds							
3)	L1 AR-1016-1	6.281	5.204	950629	604621	561.224	512.886
4)	L1 AR-1016-2	6.305	5.224	1347437	852040	558.814	524.610
5)	L1 AR-1016-3	6.371	5.414	851848	464820	554.136	515.198
6)	L1 AR-1016-4	6.478	5.468	713979	354517	555.390	499.403
7)	L1 AR-1016-5	6.793	5.695	704556	444219	550.841	490.185
8)	L2 AR-1221-1	5.208	4.209	94600	60659	168.504	155.876
9)	L2 AR-1221-2	5.311	4.308	138690	89366	321.837	299.400
10)	L2 AR-1221-3	5.397	4.396	518041	332474	401.825	368.177
11)	L3 AR-1232-1	5.397	4.396	518041	332474	486.388	451.937
12)	L3 AR-1232-2	5.986	5.224	705376	852040	1282.557	1274.228
13)	L3 AR-1232-3	6.305	5.414	1347437	464820	1319.002	1281.289
14)	L3 AR-1232-4	6.478	5.512	713979	430103	1322.719	1373.505
15)	L3 AR-1232-5	6.576	5.695	557641	444219	1521.523	1247.128
16)	L4 AR-1242-1	6.281	5.204	950629	604621	741.322	679.278
17)	L4 AR-1242-2	6.305	5.224	1347437	852040	738.395	698.503
18)	L4 AR-1242-3	6.371	5.414	851848	464820	726.114	690.791
19)	L4 AR-1242-4	6.478	5.512	713979	430103	737.760	664.325
20)	L4 AR-1242-5	7.261	6.075	119110	381967	113.235	468.152 #
21)	L5 AR-1248-1	6.281	5.204	950629	604621	942.314	867.750
22)	L5 AR-1248-2	6.576	5.468	557641	354517	410.617	374.288
23)	L5 AR-1248-3	6.793	5.512	704556	430103	433.020	434.694
24)	L5 AR-1248-4	7.221	5.695	151877	444219	83.186	379.763 #
25)	L5 AR-1248-5	7.261	6.117	119110	59320	66.060	50.009
26)	L6 AR-1254-1	7.191	6.075	572788	381967	325.695	225.434 #
27)	L6 AR-1254-2	7.420	6.235	516900	296593	191.040	199.403
28)	L6 AR-1254-3	7.802	6.674	350348	417345	118.279	171.665 #
29)	L6 AR-1254-4	8.096	6.895	239236	88081	108.825	56.919 #
30)	L6 AR-1254-5	8.526	7.324	1624326	1077836	683.023	501.252 #
31)	L7 AR-1260-1	7.969	6.792	1138631	775197	524.593	481.477
32)	L7 AR-1260-2	8.234	6.991	1373609	943266	535.949	486.404
33)	L7 AR-1260-3	8.600	7.145	1015179	884992	506.455	486.609
34)	L7 AR-1260-4	8.831	7.629	1228346	753903	518.455	486.155
35)	L7 AR-1260-5	9.159	7.878	2315691	1787842	518.991	486.521
36)	L8 AR-1262-1	8.600	7.324	1015179	1077836	356.272	904.311 #
37)	L8 AR-1262-2	9.159	7.878	2315691	1787842	455.771	456.078
38)	L8 AR-1262-3	9.494f	8.165	1595736	443517	430.750	278.688 #
39)	L8 AR-1262-4	9.547f	8.227	1045981	1336792	364.180	449.136
40)	L8 AR-1262-5	10.237	8.727	849558	497993	402.113	349.458
41)	L9 AR-1268-1	9.494f	8.165	1595736	443517	265.978	95.245 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP063021\
 Data File : PP036905.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Jun 2021 10:26
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 30 13:56:15 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 18 06:06:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.547f	8.227	1045981	1336792	183.494	309.903 #
43) L9	AR-1268-3	9.795	8.438	44416	20636	9.239	5.652 #
44) L9	AR-1268-4	10.237	8.727	849558	497993	360.670	313.061
45) L9	AR-1268-5	10.654	9.006	248568	139823	16.164	11.827 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP063021\
Data File : PP036905.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Jun 2021 10:26
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 30 13:56:15 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061721.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 18 06:06:19 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.962 min
Delta R.T.: -0.003 min
Response: 2579691
Conc: 58.37 ng/ml

Instrument :
ECD_P
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.954 min
Delta R.T.: -0.005 min
Response: 1704280
Conc: 55.02 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.994 min
Delta R.T.: -0.004 min
Response: 2408167
Conc: 56.14 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.249 min
Delta R.T.: -0.014 min
Response: 1463254
Conc: 45.75 ng/ml



#3 AR-1016-1


R.T.: 6.281 min
Delta R.T.: -0.003 min
Response: 950629
Conc: 561.22 ng/ml

Instrument :
ECD_P
ClientSampleId :




#3 AR-1016-1

R.T.: 5.204 min
Delta R.T.: -0.009 min
Response: 604621
Conc: 512.89 ng/ml




#4 AR-1016-2

R.T.: 6.305 min
Delta R.T.: -0.004 min
Response: 1347437
Conc: 558.81 ng/ml



#4 AR-1016-2


R.T.: 5.224 min
Delta R.T.: -0.009 min
Response: 852040
Conc: 524.61 ng/ml



#5 AR-1016-3


R.T.: 6.371 min
Delta R.T.: -0.004 min
Response: 851848
Conc: 554.14 ng/ml

Instrument :
ECD_P
ClientSampleId :




#5 AR-1016-3

R.T.: 5.414 min
Delta R.T.: -0.010 min
Response: 464820
Conc: 515.20 ng/ml




#6 AR-1016-4

R.T.: 6.478 min
Delta R.T.: -0.004 min
Response: 713979
Conc: 555.39 ng/ml



#6 AR-1016-4


R.T.: 5.468 min
Delta R.T.: -0.009 min
Response: 354517
Conc: 499.40 ng/ml



#7 AR-1016-5


R.T.: 6.793 min
Delta R.T.: -0.004 min
Response: 704556
Conc: 550.84 ng/ml

Instrument :
ECD_P
ClientSampleId :




#7 AR-1016-5

R.T.: 5.695 min
Delta R.T.: -0.009 min
Response: 444219
Conc: 490.19 ng/ml




#8 AR-1221-1

R.T.: 5.208 min
Delta R.T.: -0.008 min
Response: 94600
Conc: 168.50 ng/ml



#8 AR-1221-1


R.T.: 4.209 min
Delta R.T.: -0.008 min
Response: 60659
Conc: 155.88 ng/ml



#9 AR-1221-2


R.T.: 5.311 min
Delta R.T.: -0.005 min
Response: 138690
Conc: 321.84 ng/ml

Instrument :
ECD_P
ClientSampleId :




#9 AR-1221-2

R.T.: 4.308 min
Delta R.T.: -0.008 min
Response: 89366
Conc: 299.40 ng/ml




#10 AR-1221-3

R.T.: 5.397 min
Delta R.T.: -0.005 min
Response: 518041
Conc: 401.82 ng/ml



#10 AR-1221-3


R.T.: 4.396 min
Delta R.T.: -0.008 min
Response: 332474
Conc: 368.18 ng/ml



#11 AR-1232-1


R.T.: 5.397 min
Delta R.T.: -0.003 min
Response: 518041
Conc: 486.39 ng/ml

Instrument :
ECD_P
ClientSampleId :




#11 AR-1232-1

R.T.: 4.396 min
Delta R.T.: -0.007 min
Response: 332474
Conc: 451.94 ng/ml




#12 AR-1232-2

R.T.: 5.986 min
Delta R.T.: -0.003 min
Response: 705376
Conc: 1282.56 ng/ml



#12 AR-1232-2


R.T.: 5.224 min
Delta R.T.: -0.009 min
Response: 852040
Conc: 1274.23 ng/ml



#13 AR-1232-3


R.T.: 6.305 min
Delta R.T.: -0.004 min
Response: 1347437
Conc: 1319.00 ng/ml

Instrument :
ECD_P
ClientSampleId :




#13 AR-1232-3

R.T.: 5.414 min
Delta R.T.: -0.009 min
Response: 464820
Conc: 1281.29 ng/ml




#14 AR-1232-4

R.T.: 6.478 min
Delta R.T.: -0.004 min
Response: 713979
Conc: 1322.72 ng/ml



#14 AR-1232-4


R.T.: 5.512 min
Delta R.T.: -0.010 min
Response: 430103
Conc: 1373.51 ng/ml



#15 AR-1232-5


R.T.: 6.576 min
Delta R.T.: -0.003 min
Response: 557641
Conc: 1521.52 ng/ml

Instrument :
ECD_P
ClientSampleId :




#15 AR-1232-5

R.T.: 5.695 min
Delta R.T.: -0.009 min
Response: 444219
Conc: 1247.13 ng/ml




#16 AR-1242-1

R.T.: 6.281 min
Delta R.T.: -0.003 min
Response: 950629
Conc: 741.32 ng/ml



#16 AR-1242-1


R.T.: 5.204 min
Delta R.T.: -0.009 min
Response: 604621
Conc: 679.28 ng/ml



#17 AR-1242-2


R.T.: 6.305 min
Delta R.T.: -0.004 min
Response: 1347437
Conc: 738.39 ng/ml

Instrument :
ECD_P
ClientSampleId :




#17 AR-1242-2

R.T.: 5.224 min
Delta R.T.: -0.009 min
Response: 852040
Conc: 698.50 ng/ml




#18 AR-1242-3

R.T.: 6.371 min
Delta R.T.: -0.004 min
Response: 851848
Conc: 726.11 ng/ml



#18 AR-1242-3


R.T.: 5.414 min
Delta R.T.: -0.009 min
Response: 464820
Conc: 690.79 ng/ml



#19 AR-1242-4


R.T.: 6.478 min
Delta R.T.: -0.004 min
Response: 713979
Conc: 737.76 ng/ml

Instrument :
ECD_P
ClientSampleId :




#19 AR-1242-4

R.T.: 5.512 min
Delta R.T.: -0.009 min
Response: 430103
Conc: 664.32 ng/ml




#20 AR-1242-5

R.T.: 7.261 min
Delta R.T.: -0.003 min
Response: 119110
Conc: 113.24 ng/ml



#20 AR-1242-5


R.T.: 6.075 min
Delta R.T.: -0.009 min
Response: 381967
Conc: 468.15 ng/ml



#21 AR-1248-1


R.T.: 6.281 min
Delta R.T.: -0.003 min
Response: 950629
Conc: 942.31 ng/ml

Instrument :
ECD_P
ClientSampleId :




#21 AR-1248-1

R.T.: 5.204 min
Delta R.T.: -0.009 min
Response: 604621
Conc: 867.75 ng/ml




#22 AR-1248-2

R.T.: 6.576 min
Delta R.T.: -0.003 min
Response: 557641
Conc: 410.62 ng/ml



#22 AR-1248-2


R.T.: 5.468 min
Delta R.T.: -0.008 min
Response: 354517
Conc: 374.29 ng/ml



#23 AR-1248-3


R.T.: 6.793 min
Delta R.T.: -0.003 min
Response: 704556
Conc: 433.02 ng/ml

Instrument :
ECD_P
ClientSampleId :




#23 AR-1248-3

R.T.: 5.512 min
Delta R.T.: -0.008 min
Response: 430103
Conc: 434.69 ng/ml




#24 AR-1248-4

R.T.: 7.221 min
Delta R.T.: -0.003 min
Response: 151877
Conc: 83.19 ng/ml



#24 AR-1248-4


R.T.: 5.695 min
Delta R.T.: -0.009 min
Response: 444219
Conc: 379.76 ng/ml



#25 AR-1248-5


R.T.: 7.261 min
Delta R.T.: -0.004 min
Response: 119110
Conc: 66.06 ng/ml

Instrument :
ECD_P
ClientSampleId :




#25 AR-1248-5

R.T.: 6.117 min
Delta R.T.: -0.010 min
Response: 59320
Conc: 50.01 ng/ml




#26 AR-1254-1

R.T.: 7.191 min
Delta R.T.: -0.004 min
Response: 572788
Conc: 325.70 ng/ml



#26 AR-1254-1


R.T.: 6.075 min
Delta R.T.: -0.010 min
Response: 381967
Conc: 225.43 ng/ml



#27 AR-1254-2


R.T.: 7.420 min
Delta R.T.: -0.004 min
Response: 516900
Conc: 191.04 ng/ml

Instrument :
ECD_P
ClientSampleId :




#27 AR-1254-2

R.T.: 6.235 min
Delta R.T.: -0.010 min
Response: 296593
Conc: 199.40 ng/ml




#28 AR-1254-3

R.T.: 7.802 min
Delta R.T.: -0.003 min
Response: 350348
Conc: 118.28 ng/ml



#28 AR-1254-3


R.T.: 6.674 min
Delta R.T.: 0.009 min
Response: 417345
Conc: 171.66 ng/ml



#29 AR-1254-4


R.T.: 8.096 min
Delta R.T.: -0.005 min
Response: 239236
Conc: 108.83 ng/ml

Instrument :
ECD_P
ClientSampleId :




#29 AR-1254-4

R.T.: 6.895 min
Delta R.T.: -0.010 min
Response: 88081
Conc: 56.92 ng/ml




#30 AR-1254-5

R.T.: 8.526 min
Delta R.T.: -0.004 min
Response: 1624326
Conc: 683.02 ng/ml



#30 AR-1254-5


R.T.: 7.324 min
Delta R.T.: -0.011 min
Response: 1077836
Conc: 501.25 ng/ml



#31 AR-1260-1


R.T.: 7.969 min
Delta R.T.: -0.003 min
Response: 1138631
Conc: 524.59 ng/ml

Instrument :
ECD_P
ClientSampleId :




#31 AR-1260-1

R.T.: 6.792 min
Delta R.T.: -0.011 min
Response: 775197
Conc: 481.48 ng/ml




#32 AR-1260-2

R.T.: 8.234 min
Delta R.T.: -0.003 min
Response: 1373609
Conc: 535.95 ng/ml



#32 AR-1260-2


R.T.: 6.991 min
Delta R.T.: -0.010 min
Response: 943266
Conc: 486.40 ng/ml



#33 AR-1260-3


R.T.: 8.600 min
Delta R.T.: -0.003 min
Response: 1015179
Conc: 506.45 ng/ml

Instrument :
ECD_P
ClientSampleId :




#33 AR-1260-3

R.T.: 7.145 min
Delta R.T.: -0.010 min
Response: 884992
Conc: 486.61 ng/ml




#34 AR-1260-4

R.T.: 8.831 min
Delta R.T.: -0.003 min
Response: 1228346
Conc: 518.46 ng/ml



#34 AR-1260-4


R.T.: 7.629 min
Delta R.T.: -0.011 min
Response: 753903
Conc: 486.15 ng/ml



#35 AR-1260-5


R.T.: 9.159 min
Delta R.T.: -0.003 min
Response: 2315691
Conc: 518.99 ng/ml

Instrument :
ECD_P
ClientSampleId :




#35 AR-1260-5

R.T.: 7.878 min
Delta R.T.: -0.010 min
Response: 1787842
Conc: 486.52 ng/ml




#36 AR-1262-1

R.T.: 8.600 min
Delta R.T.: -0.002 min
Response: 1015179
Conc: 356.27 ng/ml



#36 AR-1262-1


R.T.: 7.324 min
Delta R.T.: -0.009 min
Response: 1077836
Conc: 904.31 ng/ml



#37 AR-1262-2


R.T.: 9.159 min
Delta R.T.: -0.002 min
Response: 2315691
Conc: 455.77 ng/ml

Instrument :
ECD_P
ClientSampleId :




#37 AR-1262-2

R.T.: 7.878 min
Delta R.T.: -0.009 min
Response: 1787842
Conc: 456.08 ng/ml




#38 AR-1262-3

R.T.: 9.494 min
Delta R.T.: 0.019 min
Response: 1595736
Conc: 430.75 ng/ml



#38 AR-1262-3


R.T.: 8.165 min
Delta R.T.: -0.010 min
Response: 443517
Conc: 278.69 ng/ml



#39 AR-1262-4


R.T.: 9.547 min
Delta R.T.: -0.019 min
Response: 1045981
Conc: 364.18 ng/ml

Instrument :
ECD_P
ClientSampleId :




#39 AR-1262-4

R.T.: 8.227 min
Delta R.T.: -0.011 min
Response: 1336792
Conc: 449.14 ng/ml




#40 AR-1262-5

R.T.: 10.237 min
Delta R.T.: -0.003 min
Response: 849558
Conc: 402.11 ng/ml



#40 AR-1262-5


R.T.: 8.727 min
Delta R.T.: -0.010 min
Response: 497993
Conc: 349.46 ng/ml



#41 AR-1268-1


R.T.: 9.494 min
Delta R.T.: 0.021 min
Response: 1595736
Conc: 265.98 ng/ml

Instrument :
ECD_P
ClientSampleId :




#41 AR-1268-1

R.T.: 8.165 min
Delta R.T.: -0.010 min
Response: 443517
Conc: 95.24 ng/ml




#42 AR-1268-2

R.T.: 9.547 min
Delta R.T.: -0.024 min
Response: 1045981
Conc: 183.49 ng/ml



#42 AR-1268-2


R.T.: 8.227 min
Delta R.T.: -0.013 min
Response: 1336792
Conc: 309.90 ng/ml



#43 AR-1268-3


R.T.: 9.795 min
Delta R.T.: -0.004 min
Response: 44416
Conc: 9.24 ng/ml

Instrument :
ECD_P
ClientSampleId :




#43 AR-1268-3

R.T.: 8.438 min
Delta R.T.: -0.010 min
Response: 20636
Conc: 5.65 ng/ml




#44 AR-1268-4

R.T.: 10.237 min
Delta R.T.: -0.002 min
Response: 849558
Conc: 360.67 ng/ml



#44 AR-1268-4


R.T.: 8.727 min
Delta R.T.: -0.009 min
Response: 497993
Conc: 313.06 ng/ml



#45 AR-1268-5

R.T.: 10.654 min
Delta R.T.: -0.004 min
Response: 248568
Conc: 16.16 ng/ml

Instrument :
ECD_P
ClientSampleId :



#45 AR-1268-5

R.T.: 9.006 min
Delta R.T.: -0.011 min
Response: 139823
Conc: 11.83 ng/ml