

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP063021\
 Data File : PP036936.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Jun 2021 21:11
 Operator : DD\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 01 00:58:16 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 18 06:06:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.964	3.955	2382976	1574696	53.915	50.840
2) SA Decachlor...	10.998	9.252	2338817	1446496	54.521	45.231
Target Compounds						
21) L5 AR-1248-1	6.283	5.206	518021	317116	513.490	455.124
22) L5 AR-1248-2	6.577	5.469	670220	425361	493.513	449.082
23) L5 AR-1248-3	6.795	5.513	809869	443762	497.745	448.499
24) L5 AR-1248-4	7.224	5.697	912270	526712	499.670	450.287
25) L5 AR-1248-5	7.263	6.119	907141	556389	503.116	469.065



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP063021\
Data File : PP036936.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Jun 2021 21:11
Operator : DD\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 01 00:58:16 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061721.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 18 06:06:19 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.964 min
Delta R.T.: -0.001 min
Response: 2382976
Conc: 53.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.955 min
Delta R.T.: -0.004 min
Response: 1574696
Conc: 50.84 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.998 min
Delta R.T.: 0.000 min
Response: 2338817
Conc: 54.52 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.252 min
Delta R.T.: -0.011 min
Response: 1446496
Conc: 45.23 ng/ml



#21 AR-1248-1


R.T.: 6.283 min
Delta R.T.: -0.002 min
Response: 518021
Conc: 513.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 5.206 min
Delta R.T.: -0.007 min
Response: 317116
Conc: 455.12 ng/ml




#22 AR-1248-2

R.T.: 6.577 min
Delta R.T.: -0.002 min
Response: 670220
Conc: 493.51 ng/ml



#22 AR-1248-2


R.T.: 5.469 min
Delta R.T.: -0.007 min
Response: 425361
Conc: 449.08 ng/ml



#23 AR-1248-3


R.T.: 6.795 min
Delta R.T.: -0.001 min
Response: 809869
Conc: 497.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.513 min
Delta R.T.: -0.007 min
Response: 443762
Conc: 448.50 ng/ml




#24 AR-1248-4

R.T.: 7.224 min
Delta R.T.: 0.000 min
Response: 912270
Conc: 499.67 ng/ml



#24 AR-1248-4


R.T.: 5.697 min
Delta R.T.: -0.007 min
Response: 526712
Conc: 450.29 ng/ml



#25 AR-1248-5

R.T.: 7.263 min
Delta R.T.: -0.002 min
Response: 907141
Conc: 503.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 6.119 min
Delta R.T.: -0.007 min
Response: 556389
Conc: 469.07 ng/ml