

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP063022\
 Data File : PP049084.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jun 2022 16:25
 Operator : YP\AJ
 Sample : N3214-03
 Misc : AR1248 MDL 25 PPB
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 MDL-SOIL-03-QT2-2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 01 04:48:06 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 29 09:37:18 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.343	3.593	26651312	48268380	18.748	18.850
2) SA Decachlor...	10.092	8.567	27054939	22998758	22.057	20.403
Target Compounds						
21) L5 AR-1248-1	5.520	4.669	713985	1319537	22.202	22.703
22) L5 AR-1248-2	5.793	4.903	1148345	1878639	24.702	27.993
23) L5 AR-1248-3	6.000	4.944	1211330	2036852	22.621	24.852
24) L5 AR-1248-4	6.408	5.116	1288683	2365392	20.852	23.978
25) L5 AR-1248-5	6.445	5.506	1266554	2284398	21.968	22.773

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP063022\
Data File : PP049084.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jun 2022 16:25
Operator : YP\AJ
Sample : N3214-03
Misc : AR1248 MDL 25 PPB
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
MDL-SOIL-03-QT2-2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 01 04:48:06 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 29 09:37:18 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

