

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070124\
 Data File : PP066165.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jul 2024 23:44
 Operator : YP\AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 02 04:42:24 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070124.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 02 04:38:24 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.780	4.018	46228961	59948732	25.064	26.067
2) SA Decachlor...	10.774	9.192	53219173	32278870	26.711	25.774
Target Compounds						
26) L6 AR-1254-1	6.821	5.950	21563956	24048274	276.975	268.537
27) L6 AR-1254-2	7.039	6.097	32549788	20749099	277.800	269.198
28) L6 AR-1254-3	7.409	6.509	32037434	30580442	267.148	264.188
29) L6 AR-1254-4	7.694	6.737	23401985	20398641	268.124	307.754
30) L6 AR-1254-5	8.115	7.160	31008639	25264936	319.823	261.315



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070124\
Data File : PP066165.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Jul 2024 23:44
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 02 04:42:24 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070124.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 02 04:38:24 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.780 min
Delta R.T.: 0.000 min
Response: 46228961
Conc: 25.06 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250




#1 Tetrachloro-m-xylene

R.T.: 4.018 min
Delta R.T.: 0.000 min
Response: 59948732
Conc: 26.07 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.774 min
Delta R.T.: 0.000 min
Response: 53219173
Conc: 26.71 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.192 min
Delta R.T.: 0.000 min
Response: 32278870
Conc: 25.77 ng/ml



#26 AR-1254-1


R.T.: 6.821 min
Delta R.T.: 0.000 min
Response: 21563956
Conc: 276.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250




#26 AR-1254-1

R.T.: 5.950 min
Delta R.T.: 0.000 min
Response: 24048274
Conc: 268.54 ng/ml




#27 AR-1254-2

R.T.: 7.039 min
Delta R.T.: 0.000 min
Response: 32549788
Conc: 277.80 ng/ml



#27 AR-1254-2


R.T.: 6.097 min
Delta R.T.: 0.000 min
Response: 20749099
Conc: 269.20 ng/ml



#28 AR-1254-3


R.T.: 7.409 min
Delta R.T.: 0.000 min
Response: 32037434
Conc: 267.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250




#28 AR-1254-3

R.T.: 6.509 min
Delta R.T.: 0.000 min
Response: 30580442
Conc: 264.19 ng/ml




#29 AR-1254-4

R.T.: 7.694 min
Delta R.T.: 0.000 min
Response: 23401985
Conc: 268.12 ng/ml



#29 AR-1254-4


R.T.: 6.737 min
Delta R.T.: 0.000 min
Response: 20398641
Conc: 307.75 ng/ml



#30 AR-1254-5

R.T.: 8.115 min
Delta R.T.: 0.000 min
Response: 31008639
Conc: 319.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250



#30 AR-1254-5

R.T.: 7.160 min
Delta R.T.: 0.000 min
Response: 25264936
Conc: 261.32 ng/ml