

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070124\
 Data File : PP066170.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jul 2024 01:06
 Operator : YP\AJ
 Sample : AR1268ICC500
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 02 05:24:36 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070124.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 02 05:21:23 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.779	4.020	90800152	114.3E6	50.000	50.000
2)	SA Decachlor...	10.774	9.192	167.7E6	105.1E6	50.000	50.000
Target Compounds							
41)	L9 AR-1268-1	9.119	7.985	126.8E6	92141495	500.000	500.000
42)	L9 AR-1268-2	9.221	8.051	122.7E6	86563461	500.000	500.000
43)	L9 AR-1268-3	9.473	8.268	102.6E6	73334314	500.000	500.000
44)	L9 AR-1268-4	9.937	8.577	39854647	32012632	500.000	500.000
45)	L9 AR-1268-5	10.397	8.903	333.3E6	218.5E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070124\
Data File : PP066170.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jul 2024 01:06
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 02 05:24:36 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070124.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 02 05:21:23 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

