

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070124\
 Data File : PP066174.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jul 2024 02:11
 Operator : YP\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP070124

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 02 02:58:11 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070124.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 02 02:53:42 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.779	4.019	93470952	116.5E6	49.926	48.938
2)	SA Decachlor...	10.773	9.191	99583757	62793696	50.759	52.363
Target Compounds							
16)	L4 AR-1242-1	5.959	5.132	27570490	27599090	501.208	494.728
17)	L4 AR-1242-2	5.982	5.152	40444414	38587475	500.117	491.082
18)	L4 AR-1242-3	6.046	5.334	25436904	20671106	506.667	489.572
19)	L4 AR-1242-4	6.145	5.416	20622244	20361267	497.586	490.151
20)	L4 AR-1242-5	6.886	5.948	22551910	22648123	501.491	490.590



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070124\
Data File : PP066174.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jul 2024 02:11
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP070124

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 02 02:58:11 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070124.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 02 02:53:42 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.779 min
Delta R.T.: -0.002 min
Response: 93470952
Conc: 49.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070124




#1 Tetrachloro-m-xylene

R.T.: 4.019 min
Delta R.T.: -0.001 min
Response: 116459305
Conc: 48.94 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.773 min
Delta R.T.: -0.005 min
Response: 99583757
Conc: 50.76 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.191 min
Delta R.T.: -0.004 min
Response: 62793696
Conc: 52.36 ng/ml



#16 AR-1242-1


R.T.: 5.959 min
Delta R.T.: -0.003 min
Response: 27570490
Conc: 501.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070124




#16 AR-1242-1

R.T.: 5.132 min
Delta R.T.: -0.002 min
Response: 27599090
Conc: 494.73 ng/ml




#17 AR-1242-2

R.T.: 5.982 min
Delta R.T.: -0.002 min
Response: 40444414
Conc: 500.12 ng/ml



#17 AR-1242-2


R.T.: 5.152 min
Delta R.T.: -0.001 min
Response: 38587475
Conc: 491.08 ng/ml



#18 AR-1242-3


R.T.: 6.046 min
Delta R.T.: -0.002 min
Response: 25436904
Conc: 506.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070124




#18 AR-1242-3

R.T.: 5.334 min
Delta R.T.: -0.002 min
Response: 20671106
Conc: 489.57 ng/ml




#19 AR-1242-4

R.T.: 6.145 min
Delta R.T.: -0.002 min
Response: 20622244
Conc: 497.59 ng/ml



#19 AR-1242-4


R.T.: 5.416 min
Delta R.T.: -0.002 min
Response: 20361267
Conc: 490.15 ng/ml



#20 AR-1242-5

R.T.: 6.886 min
Delta R.T.: -0.003 min
Response: 22551910
Conc: 501.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP070124



#20 AR-1242-5

R.T.: 5.948 min
Delta R.T.: -0.003 min
Response: 22648123
Conc: 490.59 ng/ml