

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070221\
 Data File : PP037008.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 02 Jul 2021 21:34
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 03 02:11:51 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 18 06:06:19 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.961	3.953	2510021	1612343	56.790	52.056
2) SA Decachlor...	10.995	9.248	2200222	1352908	51.290	42.304
Target Compounds						
3) L1 AR-1016-1	6.281	5.204	908823	570681	536.543	484.096
4) L1 AR-1016-2	6.305	5.224	1287719	789776	534.047	486.273
5) L1 AR-1016-3	6.370	5.414	821555	435804	534.430	483.037
6) L1 AR-1016-4	6.478	5.468	690590	333500	537.196	469.795
7) L1 AR-1016-5	6.793	5.695	681572	440388	532.872	485.958
31) L7 AR-1260-1	7.968	6.792	1160773	735744	534.794	456.973
32) L7 AR-1260-2	8.234	6.991	1366492	889038	533.172	458.441
33) L7 AR-1260-3	8.599	7.144	1028836	855613	513.268	470.455
34) L7 AR-1260-4	8.830	7.628	1225907	707293	517.426	456.098
35) L7 AR-1260-5	9.159	7.877	2304551	1654119	516.494	450.131



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070221\
Data File : PP037008.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 02 Jul 2021 21:34
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampled :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 03 02:11:51 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP061721.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 18 06:06:19 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.961 min
Delta R.T.: -0.004 min
Response: 2510021
Conc: 56.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.953 min
Delta R.T.: -0.006 min
Response: 1612343
Conc: 52.06 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.995 min
Delta R.T.: -0.003 min
Response: 2200222
Conc: 51.29 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.248 min
Delta R.T.: -0.012 min
Response: 1352908
Conc: 42.30 ng/ml



#3 AR-1016-1


R.T.: 6.281 min
Delta R.T.: -0.004 min
Response: 908823
Conc: 536.54 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 5.204 min
Delta R.T.: -0.009 min
Response: 570681
Conc: 484.10 ng/ml




#4 AR-1016-2

R.T.: 6.305 min
Delta R.T.: -0.005 min
Response: 1287719
Conc: 534.05 ng/ml



#4 AR-1016-2


R.T.: 5.224 min
Delta R.T.: -0.009 min
Response: 789776
Conc: 486.27 ng/ml



#5 AR-1016-3


R.T.: 6.370 min
Delta R.T.: -0.004 min
Response: 821555
Conc: 534.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.414 min
Delta R.T.: -0.010 min
Response: 435804
Conc: 483.04 ng/ml




#6 AR-1016-4

R.T.: 6.478 min
Delta R.T.: -0.004 min
Response: 690590
Conc: 537.20 ng/ml



#6 AR-1016-4


R.T.: 5.468 min
Delta R.T.: -0.009 min
Response: 333500
Conc: 469.80 ng/ml



#7 AR-1016-5


R.T.: 6.793 min
Delta R.T.: -0.004 min
Response: 681572
Conc: 532.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.695 min
Delta R.T.: -0.010 min
Response: 440388
Conc: 485.96 ng/ml




#31 AR-1260-1

R.T.: 7.968 min
Delta R.T.: -0.003 min
Response: 1160773
Conc: 534.79 ng/ml



#31 AR-1260-1


R.T.: 6.792 min
Delta R.T.: -0.011 min
Response: 735744
Conc: 456.97 ng/ml



#32 AR-1260-2


R.T.: 8.234 min
Delta R.T.: -0.004 min
Response: 1366492
Conc: 533.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.991 min
Delta R.T.: -0.011 min
Response: 889038
Conc: 458.44 ng/ml




#33 AR-1260-3

R.T.: 8.599 min
Delta R.T.: -0.004 min
Response: 1028836
Conc: 513.27 ng/ml



#33 AR-1260-3


R.T.: 7.144 min
Delta R.T.: -0.011 min
Response: 855613
Conc: 470.46 ng/ml



#34 AR-1260-4


R.T.: 8.830 min
Delta R.T.: -0.004 min
Response: 1225907
Conc: 517.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.628 min
Delta R.T.: -0.012 min
Response: 707293
Conc: 456.10 ng/ml



#35 AR-1260-5

R.T.: 9.159 min
Delta R.T.: -0.004 min
Response: 2304551
Conc: 516.49 ng/ml



#35 AR-1260-5

R.T.: 7.877 min
Delta R.T.: -0.011 min
Response: 1654119
Conc: 450.13 ng/ml