

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070225\
 Data File : PP073449.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jul 2025 10:27
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 03 01:34:41 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 02 04:02:48 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.490	3.781	84832653	77251520	53.144	43.558
2) SA Decachlor...	10.181	8.788	74242657	67068342	52.247	48.472
Target Compounds						
16) L4 AR-1242-1	5.642	4.861	24565175	25520347	483.465	436.988
17) L4 AR-1242-2	5.663	4.879	38441145	37166196	506.974	430.782
18) L4 AR-1242-3	5.726	5.056	23219609	20650547	483.369	450.056
19) L4 AR-1242-4	5.822	5.139	19550358	19836829	507.723	451.762
20) L4 AR-1242-5	6.552	5.662	21395713	24723159	503.384	451.840



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070225\
Data File : PP073449.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jul 2025 10:27
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 03 01:34:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 02 04:02:48 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.490 min
Delta R.T.: -0.003 min
Response: 84832653
Conc: 53.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.781 min
Delta R.T.: -0.003 min
Response: 77251520
Conc: 43.56 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.181 min
Delta R.T.: -0.005 min
Response: 74242657
Conc: 52.25 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.788 min
Delta R.T.: -0.002 min
Response: 67068342
Conc: 48.47 ng/ml



#16 AR-1242-1


R.T.: 5.642 min
Delta R.T.: -0.004 min
Response: 24565175
Conc: 483.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.861 min
Delta R.T.: -0.004 min
Response: 25520347
Conc: 436.99 ng/ml




#17 AR-1242-2

R.T.: 5.663 min
Delta R.T.: -0.004 min
Response: 38441145
Conc: 506.97 ng/ml



#17 AR-1242-2


R.T.: 4.879 min
Delta R.T.: -0.003 min
Response: 37166196
Conc: 430.78 ng/ml



#18 AR-1242-3


R.T.: 5.726 min
Delta R.T.: -0.003 min
Response: 23219609
Conc: 483.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 5.056 min
Delta R.T.: -0.003 min
Response: 20650547
Conc: 450.06 ng/ml




#19 AR-1242-4

R.T.: 5.822 min
Delta R.T.: -0.004 min
Response: 19550358
Conc: 507.72 ng/ml



#19 AR-1242-4


R.T.: 5.139 min
Delta R.T.: -0.004 min
Response: 19836829
Conc: 451.76 ng/ml



#20 AR-1242-5

R.T.: 6.552 min
Delta R.T.: -0.004 min
Response: 21395713
Conc: 503.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.662 min
Delta R.T.: -0.003 min
Response: 24723159
Conc: 451.84 ng/ml