

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070225\
 Data File : PP073466.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jul 2025 19:36
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 03 01:39:01 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 02 04:02:48 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.487	3.783	72065957	88288309	45.146	49.781
2)	SA Decachlor...	10.178	8.787	61711571	65865866	43.429	47.603
Target Compounds							
26)	L6 AR-1254-1	6.485	5.662	29281989	53092740	412.822	484.651
27)	L6 AR-1254-2	6.702	5.811	43480739	45817566	421.572	472.435
28)	L6 AR-1254-3	7.065	6.212	45963741	71855018	404.336	487.685
29)	L6 AR-1254-4	7.347	6.440	40753180	46490001	409.636	515.268 #
30)	L6 AR-1254-5	7.764	6.857	39598065	61015520	417.818	458.503



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070225\
Data File : PP073466.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jul 2025 19:36
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 03 01:39:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 02 04:02:48 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.487 min
Delta R.T.: -0.006 min
Response: 72065957
Conc: 45.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.783 min
Delta R.T.: -0.002 min
Response: 88288309
Conc: 49.78 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.178 min
Delta R.T.: -0.007 min
Response: 61711571
Conc: 43.43 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.787 min
Delta R.T.: -0.003 min
Response: 65865866
Conc: 47.60 ng/ml



#26 AR-1254-1


R.T.: 6.485 min
Delta R.T.: -0.003 min
Response: 29281989
Conc: 412.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.662 min
Delta R.T.: -0.002 min
Response: 53092740
Conc: 484.65 ng/ml




#27 AR-1254-2

R.T.: 6.702 min
Delta R.T.: -0.003 min
Response: 43480739
Conc: 421.57 ng/ml



#27 AR-1254-2


R.T.: 5.811 min
Delta R.T.: 0.000 min
Response: 45817566
Conc: 472.43 ng/ml



#28 AR-1254-3


R.T.: 7.065 min
Delta R.T.: -0.003 min
Response: 45963741
Conc: 404.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.212 min
Delta R.T.: -0.002 min
Response: 71855018
Conc: 487.69 ng/ml




#29 AR-1254-4

R.T.: 7.347 min
Delta R.T.: -0.003 min
Response: 40753180
Conc: 409.64 ng/ml



#29 AR-1254-4


R.T.: 6.440 min
Delta R.T.: -0.002 min
Response: 46490001
Conc: 515.27 ng/ml



#30 AR-1254-5

R.T.: 7.764 min
Delta R.T.: -0.003 min
Response: 39598065
Conc: 417.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.857 min
Delta R.T.: -0.001 min
Response: 61015520
Conc: 458.50 ng/ml