

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070425\
 Data File : PP073533.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Jul 2025 15:30
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 04 23:31:17 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 02 04:02:48 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.487	3.779	62127058	87676516	38.920	49.436 #
2) SA Decachlor...	10.174	8.781	48280449	60094334	33.977	43.432 #
Target Compounds						
26) L6 AR-1254-1	6.486	5.658	25156457	55727249	354.659	508.699 #
27) L6 AR-1254-2	6.701	5.806	35604505	47011120	345.207	484.742 #
28) L6 AR-1254-3	7.065	6.207	39812843	69579179	350.228	472.239 #
29) L6 AR-1254-4	7.347	6.435	32990970	44023152	331.613	487.927 #
30) L6 AR-1254-5	7.763	6.852	31252193	59367064	329.757	446.116 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070425\
Data File : PP073533.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Jul 2025 15:30
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 04 23:31:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070125.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 02 04:02:48 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.487 min
Delta R.T.: -0.006 min
Response: 62127058
Conc: 38.92 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.779 min
Delta R.T.: -0.005 min
Response: 87676516
Conc: 49.44 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.174 min
Delta R.T.: -0.012 min
Response: 48280449
Conc: 33.98 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.781 min
Delta R.T.: -0.009 min
Response: 60094334
Conc: 43.43 ng/ml



#26 AR-1254-1


R.T.: 6.486 min
Delta R.T.: -0.003 min
Response: 25156457
Conc: 354.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.658 min
Delta R.T.: -0.006 min
Response: 55727249
Conc: 508.70 ng/ml




#27 AR-1254-2

R.T.: 6.701 min
Delta R.T.: -0.004 min
Response: 35604505
Conc: 345.21 ng/ml



#27 AR-1254-2


R.T.: 5.806 min
Delta R.T.: -0.005 min
Response: 47011120
Conc: 484.74 ng/ml



#28 AR-1254-3


R.T.: 7.065 min
Delta R.T.: -0.003 min
Response: 39812843
Conc: 350.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.207 min
Delta R.T.: -0.006 min
Response: 69579179
Conc: 472.24 ng/ml




#29 AR-1254-4

R.T.: 7.347 min
Delta R.T.: -0.003 min
Response: 32990970
Conc: 331.61 ng/ml



#29 AR-1254-4


R.T.: 6.435 min
Delta R.T.: -0.007 min
Response: 44023152
Conc: 487.93 ng/ml



#30 AR-1254-5

R.T.: 7.763 min
Delta R.T.: -0.003 min
Response: 31252193
Conc: 329.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.852 min
Delta R.T.: -0.006 min
Response: 59367064
Conc: 446.12 ng/ml