

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070522\
 Data File : PP049183.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jul 2022 12:37
 Operator : YP\AJ
 Sample : PB146010BS
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB146010BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 05 23:46:36 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062822.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 29 09:37:18 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.346	3.596	30850796	50962387	21.702	19.902
2) SA Decachlor...	10.095	8.567	28269842	23519532	23.047	20.865
Target Compounds						
3) L1 AR-1016-1	5.523	4.671	27016216	41121590	513.885	450.552
4) L1 AR-1016-2	5.546	4.690	38078917	57648165	505.293	447.957
5) L1 AR-1016-3	5.607	4.865	23852562	30797898	489.970	416.201
6) L1 AR-1016-4	5.708	4.906	19901546	24491881	512.712	478.746
7) L1 AR-1016-5	6.002	5.118	20139032	31228672	527.847	418.623
31) L7 AR-1260-1	7.130	6.146	37759380	49802378	519.044	432.299
32) L7 AR-1260-2	7.389	6.333	44342325	57624798	526.855	429.617
33) L7 AR-1260-3	7.747	6.486	28058027	52151134	501.521	428.406
34) L7 AR-1260-4	7.975	6.957	34642290	34256554	495.361	406.019
35) L7 AR-1260-5	8.295	7.197	63633171	75779911	464.401	401.967



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070522\
Data File : PP049183.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jul 2022 12:37
Operator : YP\AJ
Sample : PB146010BS
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB146010BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 05 23:46:36 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP062822.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 29 09:37:18 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.346 min
Delta R.T.: 0.000 min
Response: 30850796
Conc: 21.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB146010BS




#1 Tetrachloro-m-xylene

R.T.: 3.596 min
Delta R.T.: 0.002 min
Response: 50962387
Conc: 19.90 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.095 min
Delta R.T.: 0.000 min
Response: 28269842
Conc: 23.05 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.567 min
Delta R.T.: 0.000 min
Response: 23519532
Conc: 20.87 ng/ml



#3 AR-1016-1


R.T.: 5.523 min
Delta R.T.: 0.000 min
Response: 27016216
Conc: 513.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB146010BS




#3 AR-1016-1

R.T.: 4.671 min
Delta R.T.: 0.000 min
Response: 41121590
Conc: 450.55 ng/ml




#4 AR-1016-2

R.T.: 5.546 min
Delta R.T.: 0.000 min
Response: 38078917
Conc: 505.29 ng/ml



#4 AR-1016-2


R.T.: 4.690 min
Delta R.T.: 0.000 min
Response: 57648165
Conc: 447.96 ng/ml



#5 AR-1016-3


R.T.: 5.607 min
Delta R.T.: 0.000 min
Response: 23852562
Conc: 489.97 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB146010BS




#5 AR-1016-3

R.T.: 4.865 min
Delta R.T.: 0.000 min
Response: 30797898
Conc: 416.20 ng/ml




#6 AR-1016-4

R.T.: 5.708 min
Delta R.T.: 0.000 min
Response: 19901546
Conc: 512.71 ng/ml



#6 AR-1016-4


R.T.: 4.906 min
Delta R.T.: 0.000 min
Response: 24491881
Conc: 478.75 ng/ml



#7 AR-1016-5


R.T.: 6.002 min
Delta R.T.: 0.000 min
Response: 20139032
Conc: 527.85 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB146010BS




#7 AR-1016-5

R.T.: 5.118 min
Delta R.T.: 0.000 min
Response: 31228672
Conc: 418.62 ng/ml




#31 AR-1260-1

R.T.: 7.130 min
Delta R.T.: 0.000 min
Response: 37759380
Conc: 519.04 ng/ml



#31 AR-1260-1


R.T.: 6.146 min
Delta R.T.: 0.000 min
Response: 49802378
Conc: 432.30 ng/ml



#32 AR-1260-2


R.T.: 7.389 min
Delta R.T.: 0.000 min
Response: 44342325
Conc: 526.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB146010BS




#32 AR-1260-2

R.T.: 6.333 min
Delta R.T.: 0.000 min
Response: 57624798
Conc: 429.62 ng/ml




#33 AR-1260-3

R.T.: 7.747 min
Delta R.T.: 0.000 min
Response: 28058027
Conc: 501.52 ng/ml



#33 AR-1260-3


R.T.: 6.486 min
Delta R.T.: 0.000 min
Response: 52151134
Conc: 428.41 ng/ml



#34 AR-1260-4


R.T.: 7.975 min
Delta R.T.: 0.000 min
Response: 34642290
Conc: 495.36 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB146010BS




#34 AR-1260-4

R.T.: 6.957 min
Delta R.T.: 0.000 min
Response: 34256554
Conc: 406.02 ng/ml



#35 AR-1260-5

R.T.: 8.295 min
Delta R.T.: 0.000 min
Response: 63633171
Conc: 464.40 ng/ml



#35 AR-1260-5

R.T.: 7.197 min
Delta R.T.: 0.000 min
Response: 75779911
Conc: 401.97 ng/ml