

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070524\
 Data File : PP066196.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jul 2024 16:14
 Operator : YP\AJ
 Sample : AR1248 CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248 CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 05 16:26:05 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070124.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 02 06:34:47 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.782	4.022	100.1E6	113.2E6	56.864	48.930
2) SA Decachlor...	10.781	9.197	98580993	63131892	54.532	55.497
Target Compounds						
21) L5 AR-1248-1	5.962	5.135	21691798	20398249	517.953	484.535
22) L5 AR-1248-2	6.237	5.375	31890889	26738858	500.133	470.355
23) L5 AR-1248-3	6.445	5.419	33223733	27666249	494.824	473.221
24) L5 AR-1248-4	6.850	5.595	36794008	31310797	488.043	470.053
25) L5 AR-1248-5	6.889	5.994	37332768	28010516	477.747	475.360



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP070524\
Data File : PP066196.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jul 2024 16:14
Operator : YP\AJ
Sample : AR1248 CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248 CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 05 16:26:05 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP070124.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 02 06:34:47 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.782 min
Delta R.T.: 0.000 min
Response: 100098674
Conc: 56.86 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248 CCC500




#1 Tetrachloro-m-xylene

R.T.: 4.022 min
Delta R.T.: 0.000 min
Response: 113206331
Conc: 48.93 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.781 min
Delta R.T.: 0.002 min
Response: 98580993
Conc: 54.53 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.197 min
Delta R.T.: 0.002 min
Response: 63131892
Conc: 55.50 ng/ml



#21 AR-1248-1


R.T.: 5.962 min
Delta R.T.: 0.003 min
Response: 21691798
Conc: 517.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248 CCC500




#21 AR-1248-1

R.T.: 5.135 min
Delta R.T.: 0.002 min
Response: 20398249
Conc: 484.53 ng/ml




#22 AR-1248-2

R.T.: 6.237 min
Delta R.T.: 0.002 min
Response: 31890889
Conc: 500.13 ng/ml



#22 AR-1248-2


R.T.: 5.375 min
Delta R.T.: 0.002 min
Response: 26738858
Conc: 470.35 ng/ml



#23 AR-1248-3


R.T.: 6.445 min
Delta R.T.: 0.002 min
Response: 33223733
Conc: 494.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248 CCC500




#23 AR-1248-3

R.T.: 5.419 min
Delta R.T.: 0.002 min
Response: 27666249
Conc: 473.22 ng/ml




#24 AR-1248-4

R.T.: 6.850 min
Delta R.T.: 0.002 min
Response: 36794008
Conc: 488.04 ng/ml



#24 AR-1248-4


R.T.: 5.595 min
Delta R.T.: 0.002 min
Response: 31310797
Conc: 470.05 ng/ml



#25 AR-1248-5

R.T.: 6.889 min
Delta R.T.: 0.002 min
Response: 37332768
Conc: 477.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248 CCC500



#25 AR-1248-5

R.T.: 5.994 min
Delta R.T.: 0.002 min
Response: 28010516
Conc: 475.36 ng/ml